

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092424\
 Data File : BN034178.D
 Acq On : 23 Sep 2024 23:28
 Operator : RC/JU
 Sample : P4092-09
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B36

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/25/2024
 Supervised By :mohammad ahmed 09/25/2024

Quant Time: Sep 24 01:51:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 23 19:39:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.423	152	6597	0.400	ng/ul	0.00
4) Naphthalene-d8	10.179	136	21454	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.069	164	31023	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	16.827	188	24301m	0.400	ng/ul	0.00
17) Chrysene-d12	21.046	240	14267m	0.400	ng/ul	0.00
23) Perylene-d12	23.165	264	12173m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.035	96	48789m	7.194	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.780	152	10512	0.342	ng/ul	0.00
18) Fluoranthene-d10	18.870	212	19302	0.454	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.073	88	9937	1.320	ng/ul#	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092424\
Data File : BN034178.D
Acq On : 23 Sep 2024 23:28
Operator : RC/JU
Sample : P4092-09
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0B36

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/25/2024
Supervised By :mohammad ahmed 09/25/2024

Quant Time: Sep 24 01:51:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 23 19:39:10 2024
Response via : Initial Calibration

