

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092424\
 Data File : BN034179.D
 Acq On : 24 Sep 2024 00:04
 Operator : RC/JU
 Sample : P4092-10
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B37

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/25/2024
 Supervised By :mohammad ahmed 09/25/2024

Quant Time: Sep 24 01:51:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 23 19:39:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.423	152	6878	0.400	ng/ul	0.00
4) Naphthalene-d8	10.180	136	20829	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.074	164	11463	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.831	188	20822m	0.400	ng/ul	0.00
17) Chrysene-d12	21.046	240	14182m	0.400	ng/ul	0.00
23) Perylene-d12	23.165	264	13303m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.039	96	35112	4.966	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.780	152	9452	0.316	ng/ul	0.00
18) Fluoranthene-d10	18.870	212	17689m	0.418	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.073	88	7963	1.015	ng/ul#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092424\  
Data File : BN034179.D  
Acq On    : 24 Sep 2024  00:04  
Operator  : RC/JU  
Sample    : P4092-10  
Misc      :  
ALS Vial  : 10    Sample Multiplier: 1
```

Instrument :
BNA_N
ClientSampleId :
C0B37

Manual Integrations

Reviewed By :Yogesh Patel 09/25/2024
Supervised By :mohammad ahmed 09/25/2024

Quant Time: Sep 24 01:51:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 23 19:39:10 2024
Response via : Initial Calibration

