

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
 Data File : BN007851.D
 Acq On : 25 Sep 2019 13:24
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 04:01:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	176#	0.00
2	1,4-Dioxane	0.446	0.448	-0.4	209#	0.00
3	n-Nitrosodimethylamine	0.204	0.345	-69.1#	195#	0.00
4 S	2-Fluorophenol	1.377	1.185	13.9	174#	0.00
5 S	Phenol-d6	1.962	1.522	22.4	172#	0.00
6	bis(2-Chloroethyl)ether	1.133	1.205	-6.4	212#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	179#	-0.01
8 S	Nitrobenzene-d5	0.350	0.335	4.3	176#	0.00
9	Naphthalene	1.123	1.200	-6.9	192#	-0.01
10	Hexachlorobutadiene	0.236	0.252	-6.8	189#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.674	0.729	-8.2	200#	-0.01
12	2-Methylnaphthalene	0.759	0.833	-9.7	202#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	180#	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.142	34.9#	141	-0.01
15 S	2-Fluorobiphenyl	1.647	1.650	-0.2	194#	-0.01
16	Acenaphthylene	2.121	1.958	7.7	185#	-0.02
17	Acenaphthene	1.368	1.292	5.6	186#	-0.02
18	Fluorene	1.753	1.718	2.0	193#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	171#	-0.01
20	4-Bromophenyl-phenylether	0.243	0.262	-7.8	187#	-0.02
21	Hexachlorobenzene	0.265	0.292	-10.2	190#	-0.01
22	Pentachlorophenol	0.098	0.074	24.5	149	-0.01
23	Phenanthrene	1.246	1.343	-7.8	186#	-0.01
24	Anthracene	1.125	1.153	-2.5	181#	-0.01
25 SURR	Fluoranthene-d10	1.348	1.324	1.8	169#	-0.01
26	Fluoranthene	1.549	1.555	-0.4	173#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	144	-0.01
28	Pvrene	1.364	1.554	-13.9	170#	-0.01
29 S	Terphenyl-d14	0.983	1.152	-17.2	175#	-0.01
30	Benzo(a)anthracene	1.473	1.454	1.3	148	-0.01
31	Chrysene	1.436	1.493	-4.0	153#	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	0.563	25.7#	120	-0.02
33	Indeno(1,2,3-cd)pyrene	1.797	1.380	23.2	113	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	119	-0.02
35	Benzo(b)fluoranthene	1.389	1.453	-4.6	132	-0.01
36	Benzo(k)fluoranthene	1.373	1.584	-15.4	143	-0.02
37 C	Benzo(a)pyrene	1.305	1.324	-1.5	127	-0.01
38	Dibenzo(a,h)anthracene	1.335	1.204	9.8	112	-0.02
39	Benzo(a,h,i)perylene	1.321	1.180	10.7	111	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
Data File : BN007851.D
Acq On : 25 Sep 2019 13:24
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 04:01:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			