

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
 Data File : BN007867.D
 Acq On : 25 Sep 2019 23:04
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 05:25:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	176#	0.00
2	1,4-Dioxane	0.446	0.441	1.1	205#	0.00
3	n-Nitrosodimethylamine	0.204	0.277	-35.8#	157#	0.00
4 S	2-Fluorophenol	1.377	1.266	8.1	186#	0.00
5 S	Phenol-d6	1.962	1.652	15.8	186#	0.00
6	bis(2-Chloroethyl)ether	1.133	1.336	-17.9	235#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	175#	-0.01
8 S	Nitrobenzene-d5	0.350	0.373	-6.6	192#	0.00
9	Naphthalene	1.123	1.197	-6.6	188#	-0.01
10	Hexachlorobutadiene	0.236	0.264	-11.9	194#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.674	0.699	-3.7	187#	-0.01
12	2-Methylnaphthalene	0.759	0.794	-4.6	188#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	154#	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.166	23.9	140	-0.01
15 S	2-Fluorobiphenyl	1.647	1.728	-4.9	174#	-0.01
16	Acenaphthylene	2.121	3.916	-84.6#	316#	-0.01
17	Acenaphthene	1.368	1.435	-4.9	177#	-0.01
18	Fluorene	1.753	1.903	-8.6	182#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	125	-0.01
20	4-Bromophenyl-phenylether	0.243	0.281	-15.6	146	-0.01
21	Hexachlorobenzene	0.265	0.299	-12.8	142	-0.01
22	Pentachlorophenol	0.098	0.099	-1.0	145	-0.01
23	Phenanthrene	1.246	1.330	-6.7	135	-0.01
24	Anthracene	1.125	1.267	-12.6	145	-0.01
25 SURR	Fluoranthene-d10	1.348	1.289	4.4	121	-0.02
26	Fluoranthene	1.549	1.490	3.8	121	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	102	-0.02
28	Pvrene	1.364	1.645	-20.6	128	-0.02
29 S	Terphenyl-d14	0.983	1.110	-12.9	120	-0.02
30	Benzo(a)anthracene	1.473	1.504	-2.1	109	-0.01
31	Chrysene	1.436	1.452	-1.1	106	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	0.983	-29.7#	148	-0.02
33	Indeno(1,2,3-cd)pyrene	1.797	1.784	0.7	104	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	99	-0.03
35	Benzo(b)fluoranthene	1.389	1.428	-2.8	108	-0.02
36	Benzo(k)fluoranthene	1.373	1.362	0.8	103	-0.02
37 C	Benzo(a)pyrene	1.305	1.352	-3.6	108	-0.02
38	Dibenzo(a,h)anthracene	1.335	1.316	1.4	102	-0.03
39	Benzo(a,h,i)perylene	1.321	1.437	-8.8	113	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
Data File : BN007867.D
Acq On : 25 Sep 2019 23:04
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 05:25:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			