

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
 Data File : BN007869.D
 Acq On : 26 Sep 2019 01:28
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 02:03:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	161#	0.00
2	1,4-Dioxane	0.446	0.461	-3.4	197#	0.00
3	n-Nitrosodimethylamine	0.204	0.309	-51.5#	160#	0.00
4 S	2-Fluorophenol	1.377	1.274	7.5	171#	0.00
5 S	Phenol-d6	1.962	1.600	18.5	166#	0.00
6	bis(2-Chloroethyl)ether	1.133	1.169	-3.2	188#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	149	-0.01
8 S	Nitrobenzene-d5	0.350	0.379	-8.3	166#	0.00
9	Naphthalene	1.123	1.204	-7.2	161#	-0.01
10	Hexachlorobutadiene	0.236	0.266	-12.7	167#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.674	0.676	-0.3	154#	-0.01
12	2-Methylnaphthalene	0.759	0.773	-1.8	156#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.179	17.9	121	-0.01
15 S	2-Fluorobiphenyl	1.647	1.809	-9.8	146	-0.01
16	Acenaphthylene	2.121	2.878	-35.7#	186#	-0.02
17	Acenaphthene	1.368	1.361	0.5	134	-0.01
18	Fluorene	1.753	1.730	1.3	133	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	-0.01
20	4-Bromophenyl-phenylether	0.243	0.273	-12.3	116	-0.02
21	Hexachlorobenzene	0.265	0.307	-15.8	119	-0.01
22	Pentachlorophenol	0.098	0.109	-11.2	130	-0.01
23	Phenanthrene	1.246	1.339	-7.5	110	-0.01
24	Anthracene	1.125	1.233	-9.6	115	-0.01
25 SURR	Fluoranthene-d10	1.348	1.372	-1.8	104	-0.01
26	Fluoranthene	1.549	1.636	-5.6	109	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	99	-0.02
28	Pvrene	1.364	1.488	-9.1	112	-0.01
29 S	Terphenyl-d14	0.983	1.032	-5.0	108	-0.01
30	Benzo(a)anthracene	1.473	1.543	-4.8	108	-0.01
31	Chrysene	1.436	1.546	-7.7	109	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	0.746	1.6	109	-0.02
33	Indeno(1,2,3-cd)pyrene	1.797	1.867	-3.9	105	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	100	-0.02
35	Benzo(b)fluoranthene	1.389	1.437	-3.5	110	-0.02
36	Benzo(k)fluoranthene	1.373	1.412	-2.8	107	-0.02
37 C	Benzo(a)pyrene	1.305	1.351	-3.5	109	-0.02
38	Dibenzo(a,h)anthracene	1.335	1.327	0.6	104	-0.03
39	Benzo(a,h,i)perylene	1.321	1.371	-3.8	108	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
Data File : BN007869.D
Acq On : 26 Sep 2019 01:28
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 02:03:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			