

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
 Data File : BN007880.D
 Acq On : 26 Sep 2019 08:05
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 08:43:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	156	0.00
2	1,4-Dioxane	0.400	0.399	0.3	184	0.00
3	n-Nitrosodimethylamine	0.400	0.517	-29.3#	133	0.00
4 S	2-Fluorophenol	0.400	0.357	10.8	160	0.00
5 S	Phenol-d6	0.400	0.374	6.5	161	0.00
6	bis(2-Chloroethyl)ether	0.400	0.420	-5.0	186	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	153	-0.01
8 S	Nitrobenzene-d5	0.400	0.418	-4.5	165	0.00
9	Naphthalene	0.400	0.427	-6.7	164	-0.01
10	Hexachlorobutadiene	0.400	0.445	-11.2	169	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.418	-4.5	165	-0.01
12	2-Methylnaphthalene	0.400	0.420	-5.0	165	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	139	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.333	16.8	138	-0.01
15 S	2-Fluorobiphenyl	0.400	0.408	-2.0	153	-0.01
16	Acenaphthylene	0.400	0.475	-18.7	184	-0.01
17	Acenaphthene	0.400	0.385	3.8	146	-0.01
18	Fluorene	0.400	0.396	1.0	151	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	117	-0.01
20	4-Bromophenyl-phenylether	0.400	0.450	-12.5	134	-0.01
21	Hexachlorobenzene	0.400	0.464	-16.0	137	-0.01
22	Pentachlorophenol	0.400	0.421	-5.2	142	-0.01
23	Phenanthrene	0.400	0.424	-6.0	126	-0.01
24	Anthracene	0.400	0.434	-8.5	132	-0.01
25 SURR	Fluoranthene-d10	0.400	0.398	0.5	118	-0.01
26	Fluoranthene	0.400	0.402	-0.5	119	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	106	-0.01
28	Pvrene	0.400	0.445	-11.2	122	0.00
29 S	Terphenyl-d14	0.400	0.429	-7.2	118	-0.01
30	Benzo(a)anthracene	0.400	0.424	-6.0	117	-0.01
31	Chrysene	0.400	0.423	-5.7	115	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.410	-2.5	121	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.398	0.5	108	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	104	-0.02
35	Benzo(b)fluoranthene	0.400	0.414	-3.5	114	-0.02
36	Benzo(k)fluoranthene	0.400	0.406	-1.5	110	-0.02
37 C	Benzo(a)pyrene	0.400	0.410	-2.5	112	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.394	1.5	107	-0.02
39	Benzo(a,h,i)perylene	0.400	0.406	-1.5	110	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
Data File : BN007880.D
Acq On : 26 Sep 2019 08:05
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 08:43:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				