

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
 Data File : BN007880.D
 Acq On : 26 Sep 2019 08:05
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 08:43:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	156#	0.00
2	1,4-Dioxane	0.446	0.444	0.4	184#	0.00
3	n-Nitrosodimethylamine	0.204	0.264	-29.4#	133	0.00
4 S	2-Fluorophenol	1.377	1.230	10.7	160#	0.00
5 S	Phenol-d6	1.962	1.605	18.2	161#	0.00
6	bis(2-Chloroethyl)ether	1.133	1.190	-5.0	186#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	153#	-0.01
8 S	Nitrobenzene-d5	0.350	0.366	-4.6	165#	0.00
9	Naphthalene	1.123	1.200	-6.9	164#	-0.01
10	Hexachlorobutadiene	0.236	0.263	-11.4	169#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.674	0.705	-4.6	165#	-0.01
12	2-Methylnaphthalene	0.759	0.797	-5.0	165#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	139	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.181	17.0	138	-0.01
15 S	2-Fluorobiphenyl	1.647	1.680	-2.0	153#	-0.01
16	Acenaphthylene	2.121	2.518	-18.7	184#	-0.01
17	Acenaphthene	1.368	1.315	3.9	146	-0.01
18	Fluorene	1.753	1.736	1.0	151#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	117	-0.01
20	4-Bromophenyl-phenylether	0.243	0.273	-12.3	134	-0.01
21	Hexachlorobenzene	0.265	0.308	-16.2	137	-0.01
22	Pentachlorophenol	0.098	0.103	-5.1	142	-0.01
23	Phenanthrene	1.246	1.320	-5.9	126	-0.01
24	Anthracene	1.125	1.222	-8.6	132	-0.01
25 SURR	Fluoranthene-d10	1.348	1.342	0.4	118	-0.01
26	Fluoranthene	1.549	1.557	-0.5	119	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	106	-0.01
28	Pvrene	1.364	1.517	-11.2	122	0.00
29 S	Terphenyl-d14	0.983	1.054	-7.2	118	-0.01
30	Benzo(a)anthracene	1.473	1.561	-6.0	117	-0.01
31	Chrysene	1.436	1.520	-5.8	115	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	0.776	-2.4	121	-0.02
33	Indeno(1,2,3-cd)pyrene	1.797	1.789	0.4	108	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	104	-0.02
35	Benzo(b)fluoranthene	1.389	1.436	-3.4	114	-0.02
36	Benzo(k)fluoranthene	1.373	1.392	-1.4	110	-0.02
37 C	Benzo(a)pyrene	1.305	1.338	-2.5	112	-0.02
38	Dibenzo(a,h)anthracene	1.335	1.317	1.3	107	-0.02
39	Benzo(a,h,i)perylene	1.321	1.341	-1.5	110	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092519\
Data File : BN007880.D
Acq On : 26 Sep 2019 08:05
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 08:43:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		