

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092525\
 Data File : BN037891.D
 Acq On : 25 Sep 2025 16:24
 Operator : RC/JU
 Sample : Q3141-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-184-GW-150-152

Quant Time: Sep 25 16:50:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.818	152	8047	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	25971	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	15628	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	32900	0.400	ng	0.00
29) Chrysene-d12	21.402	240	27935	0.400	ng	# 0.00
35) Perylene-d12	23.736	264	24958	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	3445	0.188	ng	0.00
5) Phenol-d6	6.973	99	3029	0.126	ng	0.00
8) Nitrobenzene-d5	8.971	82	6023	0.304	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	11269	0.312	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	3129	0.528	ng	0.00
15) 2-Fluorobiphenyl	13.094	172	18589	0.291	ng	0.00
27) Fluoranthene-d10	19.253	212	32565	0.393	ng	0.00
31) Terphenyl-d14	19.852	244	26316	0.473	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.290	88	621	0.076	ng	# 72
24) Pentachlorophenol	16.876	266	357	0.040	ng	95
25) Phenanthrene	17.260	178	3480	0.032	ng	# 90
34) Bis(2-ethylhexyl)phtha...	21.331	149	5990	0.155	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092525\
Data File : BN037891.D
Acq On : 25 Sep 2025 16:24
Operator : RC/JU
Sample : Q3141-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-150-152

Quant Time: Sep 25 16:50:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration

