




Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN092525\  
Data File : BN037902.D  
Acq On : 25 Sep 2025 23:47  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 26 00:18:41 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Sep 20 01:53:13 2025  
Response via : Initial Calibration



TIC: BN037902.D\data.ms

**(70) Pentachlorophenol (C)**




16.875min (-0.011) 30660.06 ng

response 887138

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 64.77  |
| 264.00 | 61.60  | 56.98  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN092525\  
Data File : BN037902.D  
Acq On : 25 Sep 2025 23:47  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 26 00:18:41 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Sep 20 01:53:13 2025  
Response via : Initial Calibration



TIC: BN037902.D\data.ms

**(77) Benzdine**

19.469min (-0.005) 0.00 ng

response 5355817

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.49  |
| 183.00 | 13.20  | 12.54  |
| 0.00   | 0.00   | 0.00   |