

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092623\
 Data File : BN027864.D
 Acq On : 26 Sep 2023 16:43
 Operator : MA/JU
 Sample : PB155620BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK620

Quant Time: Sep 27 03:00:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 22 03:58:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.966	152	2654	0.400	ng/ul	0.00
4) Naphthalene-d8	10.784	136	9547	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.606	164	6489	0.400	ng/ul	0.01
13) Phenanthrene-d10	17.358	188	14775	0.400	ng/ul	0.02
17) Chrysene-d12	21.545	240	13893	0.400	ng/ul	0.02
23) Perylene-d12	24.021	264	17455	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.330	96	24956	7.465	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.368	152	6108	0.422	ng/ul	0.01
18) Fluoranthene-d10	19.385	212	16703	0.394	ng/ul	0.02
Target Compounds						
					Qvalue	
12) Fluorene	15.600	166	2124	0.074	ng/ul#	15
20) Pyrene	19.752	202	2225	0.038	ng/ul#	1
26) Benzo(a)pyrene	23.860	252	5774	0.092	ng/ul#	53


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092623\
Data File : BN027864.D
Acq On : 26 Sep 2023 16:43
Operator : MA/JU
Sample : PB155620BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK620

Quant Time: Sep 27 03:00:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 22 03:58:00 2023
Response via : Initial Calibration









#12
Fluorene
Concen: 0.074 ng/ul
RT: 15.600 min Scan# 21
Delta R.T. -0.038 min
Lab File: BN027864.D
Acq: 26 Sep 2023 16:43




Instrument :
BNA_N
ClientSampleId :
SBLK620


Tgt Ion:166 Resp: 2124
Ion Ratio Lower Upper
166 100
165 9.4 78.0 117.0#
167 35.8 11.4 17.2#



#20
Pyrene
Concen: 0.038 ng/ul
RT: 19.752 min Scan# 3732
Delta R.T. -0.010 min
Lab File: BN027864.D
Acq: 26 Sep 2023 16:43

Tgt Ion:202 Resp: 2225
Ion Ratio Lower Upper
202 100
101 163.9 11.4 17.0#
100 1197.1 9.1 13.7#





#26

Benzo(a)pyrene

Concen: 0.092 ng/ul

RT: 23.860 min Scan# 50

Delta R.T. -0.040 min

Lab File: BN027864.D


Acq: 26 Sep 2023 16:43

Instrument :

BNA_N

ClientSampleId :

SBLK620



Tgt Ion:252 Resp: 5774

Ion Ratio Lower Upper

252 100

253 50.9 21.9 32.9#

125 41.5 15.5 23.3#

