

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092623\
 Data File : BN027895.D
 Acq On : 27 Sep 2023 11:55
 Operator : MA/JU
 Sample : 04410-20
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BBHH9

Quant Time: Sep 28 00:00:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 27 03:52:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.962	152	2882	0.400	ng/ul	0.00
4) Naphthalene-d8	10.779	136	9646	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.601	164	6445	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.354	188	14959	0.400	ng/ul	0.00
17) Chrysene-d12	21.539	240	14710	0.400	ng/ul	0.00
23) Perylene-d12	24.009	264	17517	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.326	96	13625	3.753	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.363	152	4896	0.335	ng/ul	0.00
18) Fluoranthene-d10	19.381	212	16271	0.362	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.363	88	2075	0.549	ng/ul	99
7) 2-Methylnaphthalene	12.434	142	450	0.026	ng/ul	95
12) Fluorene	15.651	166	663	0.023	ng/ul#	95
14) Pentachlorophenol	16.987	266	167	0.048	ng/ul#	72
15) Phenanthrene	17.396	178	2206	0.047	ng/ul#	87
16) Anthracene	17.489	178	1955	0.050	ng/ul#	84
19) Fluoranthene	19.408	202	1517	0.026	ng/ul#	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092623\
Data File : BN027895.D
Acq On : 27 Sep 2023 11:55
Operator : MA/JU
Sample : 04410-20
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BBHH9

Quant Time: Sep 28 00:00:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 27 03:52:19 2023
Response via : Initial Calibration

