

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092624\
 Data File : BN034249.D
 Acq On : 26 Sep 2024 17:56
 Operator : RC/JU
 Sample : P4164-08DL 2X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE134D2-20240919DL

Quant Time: Sep 26 18:21:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 26 14:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.394	152	4656	0.400	ng	0.00
7) Naphthalene-d8	10.159	136	12082	0.400	ng	0.01
13) Acenaphthene-d10	14.040	164	6297	0.400	ng	0.00
19) Phenanthrene-d10	16.808	188	11632	0.400	ng	0.00
29) Chrysene-d12	21.026	240	8724	0.400	ng	0.01
35) Perylene-d12	23.136	264	9213	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.040	112	945	0.072	ng	0.00
5) Phenol-d6	6.600	99	378	0.025	ng	0.00
8) Nitrobenzene-d5	8.557	82	897	0.097	ng	0.02
11) 2-Methylnaphthalene-d10	11.757	152	2383	0.134	ng	0.00
14) 2,4,6-Tribromophenol	15.555	330	160	0.056	ng	0.00
15) 2-Fluorobiphenyl	12.661	172	3590	0.140	ng	0.00
27) Fluoranthene-d10	18.848	212	5313	0.181	ng	0.00
31) Terphenyl-d14	19.470	244	3369	0.193	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.039	88	17656	3.033	ng	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092624\
Data File : BN034249.D
Acq On : 26 Sep 2024 17:56
Operator : RC/JU
Sample : P4164-08DL 2X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE134D2-20240919DL

Quant Time: Sep 26 18:21:39 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 26 14:44:02 2024
Response via : Initial Calibration

