

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092624\
 Data File : BN034257.D
 Acq On : 26 Sep 2024 22:47
 Operator : RC/JU
 Sample : P4164-18
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT180D-20240923

Quant Time: Sep 27 01:03:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 26 14:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.402	152	5185	0.400	ng	0.00
7) Naphthalene-d8	10.159	136	13406	0.400	ng	0.01
13) Acenaphthene-d10	14.051	164	6857	0.400	ng	0.00
19) Phenanthrene-d10	16.808	188	12223	0.400	ng	# 0.00
29) Chrysene-d12	21.017	240	8105	0.400	ng	# 0.00
35) Perylene-d12	23.130	264	7905	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.040	112	1723	0.117	ng	0.00
5) Phenol-d6	6.607	99	609	0.036	ng	0.00
8) Nitrobenzene-d5	8.546	82	2747	0.268	ng	0.01
11) 2-Methylnaphthalene-d10	11.757	152	6558	0.331	ng	0.00
14) 2,4,6-Tribromophenol	15.555	330	292	0.094	ng	0.00
15) 2-Fluorobiphenyl	12.661	172	9292	0.333	ng	0.00
27) Fluoranthene-d10	18.848	212	12087	0.392	ng	0.00
31) Terphenyl-d14	19.466	244	7486	0.462	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.039	88	1337	0.206	ng	# 29

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092624\
Data File : BN034257.D
Acq On : 26 Sep 2024 22:47
Operator : RC/JU
Sample : P4164-18
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT180D-20240923

Quant Time: Sep 27 01:03:20 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 26 14:44:02 2024
Response via : Initial Calibration

