

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092624\
 Data File : BN034265.D
 Acq On : 27 Sep 2024 05:32
 Operator : RC/JU
 Sample : P4164-21
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW181S-20240923

Quant Time: Sep 27 06:18:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 26 14:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.394	152	4213	0.400	ng	0.00
7) Naphthalene-d8	10.159	136	10737	0.400	ng	0.01
13) Acenaphthene-d10	14.040	164	5348	0.400	ng	0.00
19) Phenanthrene-d10	16.796	188	9681	0.400	ng	# 0.00
29) Chrysene-d12	21.017	240	7251	0.400	ng	# 0.00
35) Perylene-d12	23.130	264	6457	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.040	112	1654	0.138	ng	0.00
5) Phenol-d6	6.600	99	758	0.055	ng	0.00
8) Nitrobenzene-d5	8.547	82	1805	0.220	ng	0.01
11) 2-Methylnaphthalene-d10	11.757	152	4290	0.271	ng	0.00
14) 2,4,6-Tribromophenol	15.555	330	184	0.076	ng	0.00
15) 2-Fluorobiphenyl	12.661	172	6196	0.285	ng	0.00
27) Fluoranthene-d10	18.843	212	9382	0.384	ng	0.00
31) Terphenyl-d14	19.466	244	5924	0.409	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.039	88	1506	0.286	ng	# 35

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092624\
Data File : BN034265.D
Acq On : 27 Sep 2024 05:32
Operator : RC/JU
Sample : P4164-21
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW181S-20240923

Quant Time: Sep 27 06:18:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 26 14:44:02 2024
Response via : Initial Calibration

