

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092625\
 Data File : BN037923.D
 Acq On : 26 Sep 2025 12:50
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Sep 26 13:15:48 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.818	152	6892	0.400	ng	0.00
7) Naphthalene-d8	10.616	136	19797	0.400	ng	#-0.01
13) Acenaphthene-d10	14.473	164	10716	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	21456	0.400	ng	0.00
29) Chrysene-d12	21.402	240	18723	0.400	ng	0.00
35) Perylene-d12	23.736	264	18769	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	6509	0.414	ng	0.00
5) Phenol-d6	6.973	99	8479	0.410	ng	0.00
8) Nitrobenzene-d5	8.971	82	6335	0.420	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	10369	0.377	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	1589	0.391	ng	0.00
15) 2-Fluorobiphenyl	13.094	172	17465	0.398	ng	0.00
27) Fluoranthene-d10	19.253	212	19689	0.365	ng	0.00
31) Terphenyl-d14	19.852	244	14180	0.380	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.290	88	2894	0.414	ng	99
3) n-Nitrosodimethylamine	3.601	42	3442	0.370	ng	# 92
6) bis(2-Chloroethyl)ether	7.233	93	7646	0.399	ng	100
9) Naphthalene	10.669	128	21821	0.400	ng	99
10) Hexachlorobutadiene	10.968	225	3808	0.409	ng	# 98
12) 2-Methylnaphthalene	12.288	142	13909	0.390	ng	99
16) Acenaphthylene	14.185	152	18093	0.372	ng	100
17) Acenaphthene	14.537	154	13254	0.382	ng	99
18) Fluorene	15.523	166	15606	0.372	ng	99
20) 4,6-Dinitro-2-methylph...	15.585	198	921	0.401	ng	# 54
21) 4-Bromophenyl-phenylether	16.416	248	5194	0.371	ng	# 90
22) Hexachlorobenzene	16.528	284	6551	0.386	ng	99
23) Atrazine	16.677	200	4156	0.390	ng	# 90
24) Pentachlorophenol	16.876	266	2077	0.357	ng	96
25) Phenanthrene	17.260	178	26419	0.374	ng	100
26) Anthracene	17.347	178	22119	0.361	ng	100
28) Fluoranthene	19.285	202	28748	0.370	ng	100
30) Pyrene	19.647	202	28521	0.386	ng	100
32) Benzo(a)anthracene	21.384	228	24459	0.374	ng	100
33) Chrysene	21.438	228	27631	0.390	ng	100
34) Bis(2-ethylhexyl)phtha...	21.331	149	9722	0.376	ng	100
36) Indeno(1,2,3-cd)pyrene	26.110	276	35285	0.411	ng	98
37) Benzo(b)fluoranthene	23.031	252	30482	0.377	ng	99
38) Benzo(k)fluoranthene	23.075	252	30484	0.374	ng	99
39) Benzo(a)pyrene	23.633	252	24133	0.377	ng	98
40) Dibenzo(a,h)anthracene	26.127	278	27376	0.409	ng	99
41) Benzo(g,h,i)perylene	26.835	276	27254	0.387	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.818 min Scan# 65
Delta R.T. -0.007 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50

Instrument :
BNA_N
ClientSampleId :







Tgt Ion: 152 Resp: 6892
Ion Ratio Lower Upper
152 100
150 155.6 121.0 181.4
115 59.8 47.4 71.0



#2
1,4-Dioxane
Concen: 0.414 ng
RT: 3.290 min Scan# 29
Delta R.T. -0.007 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50

Tgt Ion: 88 Resp: 2894
Ion Ratio Lower Upper
88 100
43 31.4 26.6 39.8
58 74.2 58.9 88.3





#3

n-Nitrosodimethylamine

Concen: 0.370 ng

RT: 3.601 min Scan# 72

Delta R.T. 0.000 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Instrument :

BNA_N

ClientSampleId :






Tgt Ion: 42 Resp: 3442

Ion Ratio Lower Upper

42 100

74 120.4 93.4 140.0

44 13.4 22.2 33.4#



#4

2-Fluorophenol

Concen: 0.414 ng

RT: 5.384 min Scan# 319

Delta R.T. 0.000 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Tgt Ion: 112 Resp: 6509


Ion Ratio Lower Upper

112 100

64 55.4 44.6 66.8

63 28.9 24.9 37.3





#5

Phenol-d6

Concen: 0.410 ng

RT: 6.973 min Scan# 539

Delta R.T. 0.000 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Instrument :

BNA_N

ClientSampleId :



Tgt Ion	Ratio	Lower	Upper
99	100		
42	20.8	17.2	25.8
71	32.4	27.3	40.9



#6

bis(2-Chloroethyl)ether

Concen: 0.399 ng




RT: 7.233 min Scan# 575


Delta R.T. -0.007 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Tgt Ion	Ratio	Lower	Upper
93	100		
63	75.3	60.1	90.1
95	32.2	25.4	38.2





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.616 min Scan# 941

Delta R.T. -0.011 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50



Instrument :

BNA_N

ClientSampleId :



Tgt Ion:	136	Resp:	19797
Ion Ratio	Lower	Upper	
136	100		
137	11.1	9.0	13.4
54	9.5	5.8	8.8#
68	7.3	4.3	6.5#



#8

Nitrobenzene-d5

Concen: 0.420 ng




RT: 8.971 min Scan# 786


Delta R.T. 0.000 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Tgt Ion:	82	Resp:	6335
Ion Ratio	Lower	Upper	
82	100		
128	48.0	34.8	52.2
54	48.8	42.2	63.2









#9
 Naphthalene
 Concen: 0.400 ng
 RT: 10.669 min Scan# 945
 Delta R.T. 0.000 min
 Lab File: BN037923.D
 Acq: 26 Sep 2025 12:50




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:	128	Resp:	21821
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.2	13.8
127	13.3	11.0	16.6



#10
 Hexachlorobutadiene
 Concen: 0.409 ng
 RT: 10.968 min Scan# 973
 Delta R.T. -0.011 min
 Lab File: BN037923.D
 Acq: 26 Sep 2025 12:50

Tgt Ion:	225	Resp:	3808
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.6	51.4	77.2









#11
2-Methylnaphthalene-d10
Concen: 0.377 ng
RT: 12.217 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:152 Resp: 10369
Ion Ratio Lower Upper
152 100
151 21.7 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.390 ng
RT: 12.288 min Scan# 1177
Delta R.T. -0.005 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50

Tgt Ion:142 Resp: 13909
Ion Ratio Lower Upper
142 100
141 89.7 71.2 106.8
115 36.0 29.7 44.5





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.473 min Scan# 14

Delta R.T. 0.000 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Instrument :

BNA_N

ClientSampleId :



Tgt Ion:	164	Resp:	10716
Ion	Ratio	Lower	Upper
164	100		
162	105.5	84.8	127.2
160	55.8	46.1	69.1



#14

2,4,6-Tribromophenol

Concen: 0.391 ng




RT: 15.969 min Scan# 1590


Delta R.T. 0.000 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Tgt Ion:	330	Resp:	1589
Ion	Ratio	Lower	Upper
330	100		
332	97.8	77.2	115.8
141	45.1	37.2	55.8





#15

2-Fluorobiphenyl

Concen: 0.398 ng

RT: 13.094 min Scan# 13

Delta R.T. 0.000 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Instrument :

BNA_N

ClientSampleId :



Tgt Ion:	172	Resp:	17465
Ion Ratio	Lower	Upper	
172	100		
171	34.7	28.6	43.0
170	23.7	19.7	29.5



#16

Acenaphthylene

Concen: 0.372 ng




RT: 14.185 min Scan# 1429


Delta R.T. -0.011 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Tgt Ion:	152	Resp:	18093
Ion Ratio	Lower	Upper	
152	100		
151	19.8	15.8	23.8
153	13.2	10.5	15.7









#17
Acenaphthene
Concen: 0.382 ng
RT: 14.537 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:	154	Resp:	13254
Ion Ratio	Lower	Upper	
154	100		
153	112.1	90.5	135.7
152	63.8	51.8	77.8



#18
Fluorene
Concen: 0.372 ng
RT: 15.523 min Scan# 1554
Delta R.T. 0.000 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50

Tgt Ion:	166	Resp:	15606
Ion Ratio	Lower	Upper	
166	100		
165	100.2	79.0	118.4
167	13.6	10.6	16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 16

Delta R.T. 0.000 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50



Instrument :

BNA_N

ClientSampleId :



Tgt Ion:	188	Resp:	21456
Ion Ratio	100	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	8.3	7.6	11.4



#20

4,6-Dinitro-2-methylphenol

Concen: 0.401 ng




RT: 15.585 min Scan# 1559


Delta R.T. -0.012 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Tgt Ion:	198	Resp:	921
Ion Ratio	100	Lower	Upper
198	100		
51	118.4	59.1	88.7#
105	76.5	41.3	61.9#









#21
4-Bromophenyl-phenylether
Concen: 0.371 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:248 Resp: 5194
Ion Ratio Lower Upper
248 100
250 102.0 77.6 116.4
141 46.8 48.8 73.2#



#22
Hexachlorobenzene
Concen: 0.386 ng
RT: 16.528 min Scan# 1635
Delta R.T. -0.012 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50

Tgt Ion:284 Resp: 6551
Ion Ratio Lower Upper
284 100
142 38.1 30.9 46.3
249 30.3 23.9 35.9





#23

Atrazine

Concen: 0.390 ng

RT: 16.677 min Scan# 1647

Delta R.T. -0.012 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Instrument :

BNA_N

ClientSampleId :



Tgt Ion:	200	Resp:	4156
Ion Ratio	Lower	Upper	
200	100		
173	31.5	18.3	27.5#
215	44.4	38.6	58.0



#24

Pentachlorophenol

Concen: 0.357 ng




RT: 16.876 min Scan# 1663


Delta R.T. 0.001 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Tgt Ion:	266	Resp:	2077
Ion Ratio	Lower	Upper	
266	100		
264	62.8	50.9	76.3
268	63.0	54.3	81.5





#25

Phenanthrene

Concen: 0.374 ng

RT: 17.260 min Scan# 16

Delta R.T. 0.000 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Instrument :

BNA_N

ClientSampleId :



Tgt Ion:	178	Resp:	26419
Ion Ratio	Lower	Upper	
178	100		
176	19.4	15.7	23.5
179	15.4	12.3	18.5



#26

Anthracene

Concen: 0.361 ng




RT: 17.347 min Scan# 1701


Delta R.T. -0.012 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Tgt Ion:	178	Resp:	22119
Ion Ratio	Lower	Upper	
178	100		
176	18.7	15.0	22.4
179	15.3	12.3	18.5





#27

Fluoranthene-d10

Concen: 0.365 ng

RT: 19.253 min Scan# 19

Delta R.T. 0.000 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:212 Resp: 19689

Ion Ratio Lower Upper

212 100

106 16.1 12.6 19.0

104 9.1 7.4 11.0



#28

Fluoranthene

Concen: 0.370 ng

RT: 19.285 min Scan# 1961

Delta R.T. 0.000 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Tgt Ion:202 Resp: 28748


Ion Ratio Lower Upper

202 100

101 11.9 9.3 13.9

203 17.0 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.402 min Scan# 2272

Delta R.T. 0.000 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:240 Resp: 18723

Ion Ratio Lower Upper

240 100

120 16.8 11.4 17.2

236 29.0 23.0 34.6



#30

Pyrene

Concen: 0.386 ng

RT: 19.647 min Scan# 2039

Delta R.T. 0.000 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Tgt Ion:202 Resp: 28521


Ion Ratio Lower Upper

202 100

200 20.9 16.7 25.1


203 17.8 14.2 21.4








#31
 Terphenyl-d14
 Concen: 0.380 ng
 RT: 19.852 min Scan# 2083
 Delta R.T. 0.000 min
 Lab File: BN037923.D
 Acq: 26 Sep 2025 12:50

Instrument :
 BNA_N
 ClientSampleId :







Tgt Ion:	244	Resp:	14180
Ion	Ratio	Lower	Upper
244	100		
212	9.4	7.6	11.4
122	16.8	13.9	20.9



#32
 Benzo(a)anthracene
 Concen: 0.374 ng
 RT: 21.384 min Scan# 2270
 Delta R.T. 0.000 min
 Lab File: BN037923.D
 Acq: 26 Sep 2025 12:50

Tgt Ion:	228	Resp:	24459
Ion	Ratio	Lower	Upper
228	100		
226	28.3	22.6	33.8
229	19.3	15.6	23.4





#33

Chrysene

Concen: 0.390 ng

RT: 21.438 min Scan# 2276

Delta R.T. 0.000 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:228 Resp: 27631

Ion Ratio Lower Upper

228 100

226 30.9 24.6 37.0

229 19.1 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.376 ng

RT: 21.331 min Scan# 2264

Delta R.T. 0.009 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Tgt Ion:149 Resp: 9722


Ion Ratio Lower Upper

149 100

167 25.5 20.4 30.6

279 3.1 2.4 3.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.736 min Scan# 2836

Delta R.T. 0.009 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Instrument :

BNA_N

ClientSampleId :



Tgt Ion:	264	Resp:	18769
Ion Ratio	Lower	Upper	
264	100		
260	26.9	21.5	32.3
265	61.7	53.8	80.6



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.411 ng




RT: 26.110 min Scan# 3651


Delta R.T. 0.003 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50


Tgt Ion:	276	Resp:	35285
Ion Ratio	Lower	Upper	
276	100		
138	28.8	21.4	32.0
277	25.3	20.3	30.5







#37
Benzo(b)fluoranthene
Concen: 0.377 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.006 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50

Instrument :
BNA_N
ClientSampleId :







Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.1	19.4	29.0
125	17.4	13.0	19.4



#38
Benzo(k)fluoranthene
Concen: 0.374 ng
RT: 23.075 min Scan# 2613
Delta R.T. 0.003 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.8	19.5	29.3
125	16.0	13.0	19.4





#39

Benzo(a)pyrene

Concen: 0.377 ng

RT: 23.633 min Scan# 28

Delta R.T. 0.009 min


Lab File: BN037923.D

Acq: 26 Sep 2025 12:50




Instrument :

BNA_N

ClientSampleId :



Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.6	20.6	30.8
125	23.2	16.7	25.1



#40

Dibenzo(a,h)anthracene

Concen: 0.409 ng




RT: 26.127 min Scan# 3657


Delta R.T. 0.009 min

Lab File: BN037923.D

Acq: 26 Sep 2025 12:50


Tgt Ion	Ratio	Lower	Upper
278	100		
139	19.4	15.4	23.2
279	25.6	20.8	31.2





#41
Benzo(g,h,i)perylene
Concen: 0.387 ng
RT: 26.835 min Scan# 38
Delta R.T. 0.006 min
Lab File: BN037923.D
Acq: 26 Sep 2025 12:50

Instrument :
BNA_N
ClientSampleId :



Tgt Ion:	276	Resp:	27254
Ion	Ratio	Lower	Upper
276	100		
277	24.8	20.2	30.4
138	24.1	19.8	29.8

