




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092625\
Data File : BN037922.D
Acq On : 26 Sep 2025 12:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 29 10:38:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 29 10:34:59 2025
Response via : Initial Calibration



TIC: BN037922.D\data.ms

(70) Pentachlorophenol (C)




16.881min (-0.005) 28751.81 ng

response 1019202

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.83
264.00	61.60	64.31
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092625\
Data File : BN037922.D
Acq On : 26 Sep 2025 12:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 29 10:38:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 29 10:34:59 2025
Response via : Initial Calibration



TIC: BN037922.D\data.ms

(77) Benzidine

19.475min (+ 0.001) 0.00 ng

response 5937981

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.80
183.00	13.20	12.14
0.00	0.00	0.00