

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092719\
 Data File : BN007882.D
 Acq On : 27 Sep 2019 11:22
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 27 15:51:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	136	0.00
2	1,4-Dioxane	0.446	0.460	-3.1	166#	0.00
3	n-Nitrosodimethylamine	0.204	0.321	-57.4#	141	0.00
4 S	2-Fluorophenol	1.377	1.262	8.4	144	0.00
5 S	Phenol-d6	1.962	1.605	18.2	141	0.00
6	bis(2-Chloroethyl)ether	1.133	1.193	-5.3	163#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	135	-0.01
8 S	Nitrobenzene-d5	0.350	0.369	-5.4	147	0.00
9	Naphthalene	1.123	1.195	-6.4	145	-0.01
10	Hexachlorobutadiene	0.236	0.262	-11.0	149	-0.01
11 SURR	2-Methylnaphthalene-d10	0.674	0.707	-4.9	147	-0.01
12	2-Methylnaphthalene	0.759	0.810	-6.7	148	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	131	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.189	13.3	136	-0.01
15 S	2-Fluorobiphenyl	1.647	1.665	-1.1	142	-0.01
16	Acenaphthylene	2.121	2.057	3.0	141	-0.01
17	Acenaphthene	1.368	1.320	3.5	138	-0.01
18	Fluorene	1.753	1.696	3.3	138	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	-0.01
20	4-Bromophenyl-phenylether	0.243	0.271	-11.5	134	-0.01
21	Hexachlorobenzene	0.265	0.299	-12.8	134	-0.01
22	Pentachlorophenol	0.098	0.098	0.0	136	-0.01
23	Phenanthrene	1.246	1.341	-7.6	129	-0.01
24	Anthracene	1.125	1.200	-6.7	130	-0.01
25 SURR	Fluoranthene-d10	1.348	1.351	-0.2	119	-0.01
26	Fluoranthene	1.549	1.578	-1.9	122	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	109	-0.01
28	Pvrene	1.364	1.475	-8.1	122	-0.01
29 S	Terphenyl-d14	0.983	1.079	-9.8	124	-0.01
30	Benzo(a)anthracene	1.473	1.519	-3.1	117	-0.01
31	Chrysene	1.436	1.498	-4.3	117	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	0.787	-3.8	126	-0.02
33	Indeno(1,2,3-cd)pyrene	1.797	1.792	0.3	111	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	107	-0.02
35	Benzo(b)fluoranthene	1.389	1.421	-2.3	116	-0.01
36	Benzo(k)fluoranthene	1.373	1.433	-4.4	116	-0.02
37 C	Benzo(a)pyrene	1.305	1.338	-2.5	115	-0.02
38	Dibenzo(a,h)anthracene	1.335	1.334	0.1	112	-0.02
39	Benzo(a,h,i)perylene	1.321	1.313	0.6	111	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092719\
Data File : BN007882.D
Acq On : 27 Sep 2019 11:22
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 27 15:51:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			