

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092719\
 Data File : BN007888.D
 Acq On : 27 Sep 2019 16:40
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 27 18:45:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	166#	-0.02
2	1,4-Dioxane	0.446	0.507	-13.7	223#	-0.02
3	n-Nitrosodimethylamine	0.204	0.361	-77.0#	193#	-0.02
4 S	2-Fluorophenol	1.377	1.307	5.1	181#	-0.02
5 S	Phenol-d6	1.962	1.747	11.0	186#	-0.02
6	bis(2-Chloroethyl)ether	1.133	1.464	-29.2#	243#	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	168#	-0.01
8 S	Nitrobenzene-d5	0.350	0.390	-11.4	193#	-0.01
9	Naphthalene	1.123	1.199	-6.8	180#	-0.03
10	Hexachlorobutadiene	0.236	0.260	-10.2	184#	-0.03
11 SURR	2-Methylnaphthalene-d10	0.674	0.714	-5.9	184#	-0.02
12	2-Methylnaphthalene	0.759	0.815	-7.4	186#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	161#	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.211	3.2	186#	-0.01
15 S	2-Fluorobiphenyl	1.647	1.673	-1.6	176#	-0.01
16	Acenaphthylene	2.121	2.272	-7.1	191#	-0.02
17	Acenaphthene	1.368	1.358	0.7	174#	-0.02
18	Fluorene	1.753	1.774	-1.2	177#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	146	-0.01
20	4-Bromophenyl-phenylether	0.243	0.265	-9.1	162#	-0.02
21	Hexachlorobenzene	0.265	0.293	-10.6	163#	-0.01
22	Pentachlorophenol	0.098	0.131	-33.7#	224#	-0.02
23	Phenanthrene	1.246	1.346	-8.0	160#	-0.01
24	Anthracene	1.125	1.270	-12.9	171#	-0.01
25 SURR	Fluoranthene-d10	1.348	1.407	-4.4	154#	-0.01
26	Fluoranthene	1.549	1.631	-5.3	156#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	137	-0.01
28	Pvrene	1.364	1.538	-12.8	160#	-0.01
29 S	Terphenyl-d14	0.983	1.101	-12.0	159#	-0.01
30	Benzo(a)anthracene	1.473	1.585	-7.6	153#	-0.01
31	Chrysene	1.436	1.513	-5.4	148	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	1.389	-83.2#	280#	-0.02
33	Indeno(1,2,3-cd)pyrene	1.797	1.758	2.2	137	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	133	-0.01
35	Benzo(b)fluoranthene	1.389	1.490	-7.3	151#	-0.01
36	Benzo(k)fluoranthene	1.373	1.414	-3.0	142	-0.01
37 C	Benzo(a)pyrene	1.305	1.373	-5.2	147	-0.01
38	Dibenzo(a,h)anthracene	1.335	1.305	2.2	136	-0.02
39	Benzo(a,h,i)perylene	1.321	1.342	-1.6	140	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092719\
Data File : BN007888.D
Acq On : 27 Sep 2019 16:40
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 27 18:45:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			