

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
 Data File : BN027917.D
 Acq On : 28 Sep 2023 09:28
 Operator : MA/JU
 Sample : PB155752BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK752

Quant Time: Sep 28 10:46:52 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 27 03:52:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.954	152	3571	0.400	ng/ul	0.00
4) Naphthalene-d8	10.773	136	13504	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.596	164	9122	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.350	188	20055	0.400	ng/ul	0.00
17) Chrysene-d12	21.536	240	19023	0.400	ng/ul	0.00
23) Perylene-d12	24.006	264	22426	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.330	96	34451	7.659	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.357	152	9171	0.448	ng/ul	0.00
18) Fluoranthene-d10	19.376	212	23493	0.404	ng/ul	0.00
Target Compounds						
					Qvalue	
12) Fluorene	15.591	166	3075	0.076	ng/ul#	15
20) Pyrene	19.748	202	3045	0.038	ng/ul#	1
26) Benzo(a)pyrene	23.851	252	7911	0.098	ng/ul#	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
Data File : BN027917.D
Acq On : 28 Sep 2023 09:28
Operator : MA/JU
Sample : PB155752BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK752

Quant Time: Sep 28 10:46:52 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 27 03:52:19 2023
Response via : Initial Calibration

