

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
 Data File : BN027976.D
 Acq On : 30 Sep 2023 00:51
 Operator : MA/JU
 Sample : 04497-18
 Misc :
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BBHM6

Quant Time: Sep 30 01:43:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 27 03:52:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	6288	0.400	ng/ul	0.00
4) Naphthalene-d8	10.779	136	23291	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.601	164	13967	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.354	188	30297	0.400	ng/ul	0.00
17) Chrysene-d12	21.539	240	20836	0.400	ng/ul	0.00
23) Perylene-d12	24.018	264	20031	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.330	96	11113	1.403	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.357	152	5810	0.165	ng/ul	0.00
18) Fluoranthene-d10	19.381	212	12631	0.198	ng/ul	0.00
Target Compounds						
5) Naphthalene	10.828	128	2470	0.037	ng/ul#	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
Data File : BN027976.D
Acq On : 30 Sep 2023 00:51
Operator : MA/JU
Sample : 04497-18
Misc :
ALS Vial : 64 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BBHM6

Quant Time: Sep 30 01:43:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 27 03:52:19 2023
Response via : Initial Calibration

