

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
 Data File : BN027979.D
 Acq On : 30 Sep 2023 02:42
 Operator : MA/JU
 Sample : 04497-21
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BBJ05

Quant Time: Sep 30 03:11:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 27 03:52:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	8485	0.400	ng/ul	0.00
4) Naphthalene-d8	10.779	136	31307	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.601	164	19171	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.354	188	39935	0.400	ng/ul	0.00
17) Chrysene-d12	21.539	240	26338	0.400	ng/ul	0.00
23) Perylene-d12	24.015	264	26011	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.330	96	20010	1.872	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.357	152	12940	0.273	ng/ul	0.00
18) Fluoranthene-d10	19.380	212	27213	0.338	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.368	88	1402	0.126	ng/ul#	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
Data File : BN027979.D
Acq On : 30 Sep 2023 02:42
Operator : MA/JU
Sample : 04497-21
Misc :
ALS Vial : 67 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BBJ05

Quant Time: Sep 30 03:11:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 27 03:52:19 2023
Response via : Initial Calibration

