

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
 Data File : BN027937.D
 Acq On : 28 Sep 2023 21:55
 Operator : MA/JU
 Sample : 04480-22
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BBJE7

Quant Time: Sep 29 04:00:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 27 03:52:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	4023	0.400	ng/ul	0.00
4) Naphthalene-d8	10.779	136	15879	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.601	164	10930	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.354	188	24425	0.400	ng/ul	0.00
17) Chrysene-d12	21.536	240	21712	0.400	ng/ul	0.00
23) Perylene-d12	24.009	264	20424	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.325	96	27648	5.456	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.357	152	10851	0.451	ng/ul	0.00
18) Fluoranthene-d10	19.376	212	31398	0.473	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.368	88	1220	0.231	ng/ul#	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
Data File : BN027937.D
Acq On : 28 Sep 2023 21:55
Operator : MA/JU
Sample : 04480-22
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BBJE7

Quant Time: Sep 29 04:00:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 27 03:52:19 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.231 ng/ul
 RT: 3.368 min Scan# 115
 Delta R.T. 0.000 min
 Lab File: BN027937.D
 Acq: 28 Sep 2023 21:55

Instrument :
 BNA_N
 ClientSampleId :
 BBJE7

Tgt Ion: 88 Resp: 1220

Ion	Ratio	Lower	Upper
88	100		
43	64.4	39.6	59.4#
58	55.1	48.3	72.5

