

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
 Data File : BN027969.D
 Acq On : 29 Sep 2023 20:00
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4281

Quant Time: Sep 30 01:07:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 27 03:52:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	6932	0.400	ng/ul	0.00
4) Naphthalene-d8	10.779	136	25500	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.601	164	14825	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.358	188	30108	0.400	ng/ul	0.00
17) Chrysene-d12	21.545	240	20783	0.400	ng/ul	0.00
23) Perylene-d12	24.024	264	21091	0.400	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.330	96	3128	0.358	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.357	152	17053	0.441	ng/ul	0.00
18) Fluoranthene-d10	19.385	212	32587	0.513	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.368	88	3158	0.348	ng/ul	96
5) Naphthalene	10.828	128	29109	0.399	ng/ul	100
7) 2-Methylnaphthalene	12.434	142	18882	0.408	ng/ul	100
8) 1-Methylnaphthalene	12.654	142	19236	0.403	ng/ul	98
10) Acenaphthylene	14.333	152	28537	0.441	ng/ul	99
11) Acenaphthene	14.666	153	21031	0.376	ng/ul	97
12) Fluorene	15.651	166	23719	0.360	ng/ul	99
14) Pentachlorophenol	16.991	266	2965	0.421	ng/ul	99
15) Phenanthrene	17.401	178	36207	0.383	ng/ul	99
16) Anthracene	17.494	178	34699	0.438	ng/ul	99
19) Fluoranthene	19.413	202	39151	0.466	ng/ul	98
20) Pyrene	19.780	202	39489	0.456	ng/ul	99
21) Benzo(a)anthracene	21.527	228	31947	0.397	ng/ul	98
22) Chrysene	21.583	228	30128	0.363	ng/ul	99
24) Benzo(b)fluoranthene	23.255	252	29381	0.313	ng/ul	92
25) Benzo(k)fluoranthene	23.302	252	29397	0.323	ng/ul#	92
26) Benzo(a)pyrene	23.910	252	25920	0.343	ng/ul	92
27) Indeno(1,2,3-cd)pyrene	26.633	276	34073	0.352	ng/ul#	92
28) Dibenzo(a,h)anthracene	26.643	278	27910	0.358	ng/ul	93
29) Benzo(g,h,i)perylene	27.448	276	27975	0.340	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092823\
Data File : BN027969.D
Acq On : 29 Sep 2023 20:00
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4281

Quant Time: Sep 30 01:07:02 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 27 03:52:19 2023
Response via : Initial Calibration

