

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN093019\
 Data File : BN007892.D
 Acq On : 30 Sep 2019 12:18
 Operator : JU
 Sample : PB123468BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB123468BL

Quant Time: Oct 01 07:07:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	7605	0.40	ng	0.00
7) Naphthalene-d8	10.46	136	29122	0.40	ng	-0.01
13) Acenaphthene-d10	14.31	164	16629	0.40	ng	-0.01
19) Phenanthrene-d10	17.06	188	32470	0.40	ng	-0.01
27) Chrysene-d12	21.26	240	31390	0.40	ng	-0.01
34) Perylene-d12	23.47	264	37786	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.31	112	3885	0.15	ng	0.00
5) Phenol-d6	6.87	99	2883	0.08	ng	0.00
8) Nitrobenzene-d5	8.83	82	8878	0.35	ng	0.00
11) 2-Methylnaphthalene-d10	12.05	152	16304	0.33	ng	-0.01
14) 2,4,6-Tribromophenol	15.81	330	2127	0.23	ng	-0.01
15) 2-Fluorobiphenyl	12.93	172	23025	0.34	ng	-0.01
25) Fluoranthene-d10	19.09	212	36462	0.33	ng	-0.01
29) Terphenyl-d14	19.70	244	29859	0.39	ng	-0.01
Target Compounds						Qvalue
23) Phenanthrene	17.10	178	2966	0.029	ng	98
32) Bis(2-ethylhexyl)phthalate	21.19	149	1771	0.030	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN093019\
Data File : BN007892.D
Acq On : 30 Sep 2019 12:18
Operator : JU
Sample : PB123468BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB123468BL

Quant Time: Oct 01 07:07:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

