

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN093019\
 Data File : BN007890.D
 Acq On : 30 Sep 2019 11:06
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 01 07:01:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	170	0.00
2	1,4-Dioxane	0.400	0.413	-3.2	207	0.00
3	n-Nitrosodimethylamine	0.400	0.818	-104.5#	229	0.00
4 S	2-Fluorophenol	0.400	0.365	8.8	178	0.00
5 S	Phenol-d6	0.400	0.381	4.8	179	0.00
6	bis(2-Chloroethyl)ether	0.400	0.431	-7.7	208	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	171	-0.01
8 S	Nitrobenzene-d5	0.400	0.423	-5.7	187	-0.01
9	Naphthalene	0.400	0.426	-6.5	184	-0.01
10	Hexachlorobutadiene	0.400	0.428	-7.0	182	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.421	-5.2	186	-0.01
12	2-Methylnaphthalene	0.400	0.427	-6.7	188	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	163	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.355	11.3	173	-0.02
15 S	2-Fluorobiphenyl	0.400	0.401	-0.3	177	0.00
16	Acenaphthylene	0.400	0.398	0.5	180	-0.02
17	Acenaphthene	0.400	0.391	2.3	174	-0.02
18	Fluorene	0.400	0.392	2.0	174	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	155	-0.01
20	4-Bromophenyl-phenylether	0.400	0.426	-6.5	167	-0.01
21	Hexachlorobenzene	0.400	0.440	-10.0	172	-0.01
22	Pentachlorophenol	0.400	0.431	-7.7	191	-0.01
23	Phenanthrene	0.400	0.425	-6.2	166	-0.01
24	Anthracene	0.400	0.429	-7.2	172	0.00
25 SURR	Fluoranthene-d10	0.400	0.409	-2.2	160	-0.01
26	Fluoranthene	0.400	0.412	-3.0	161	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	149	0.00
28	Pvrene	0.400	0.421	-5.2	162	-0.01
29 S	Terphenyl-d14	0.400	0.424	-6.0	163	-0.01
30	Benzo(a)anthracene	0.400	0.411	-2.7	159	0.00
31	Chrysene	0.400	0.417	-4.2	159	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.369	7.8	153	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.396	1.0	151	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	145	-0.02
35	Benzo(b)fluoranthene	0.400	0.412	-3.0	158	-0.02
36	Benzo(k)fluoranthene	0.400	0.409	-2.2	154	-0.02
37 C	Benzo(a)pyrene	0.400	0.405	-1.3	154	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.394	1.5	149	-0.02
39	Benzo(a,h,i)perylene	0.400	0.399	0.3	150	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN093019\
Data File : BN007890.D
Acq On : 30 Sep 2019 11:06
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 01 07:01:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				