

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN093019\
 Data File : BN007890.D
 Acq On : 30 Sep 2019 11:06
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 01 07:01:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	170#	0.00
2	1,4-Dioxane	0.446	0.460	-3.1	207#	0.00
3	n-Nitrosodimethylamine	0.204	0.418	-104.9#	229#	0.00
4 S	2-Fluorophenol	1.377	1.255	8.9	178#	0.00
5 S	Phenol-d6	1.962	1.638	16.5	179#	0.00
6	bis(2-Chloroethyl)ether	1.133	1.221	-7.8	208#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	171#	-0.01
8 S	Nitrobenzene-d5	0.350	0.370	-5.7	187#	-0.01
9	Naphthalene	1.123	1.196	-6.5	184#	-0.01
10	Hexachlorobutadiene	0.236	0.253	-7.2	182#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.674	0.709	-5.2	186#	-0.01
12	2-Methylnaphthalene	0.759	0.810	-6.7	188#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	163#	-0.02
14 S	2,4,6-Tribromophenol	0.218	0.193	11.5	173#	-0.02
15 S	2-Fluorobiphenyl	1.647	1.653	-0.4	177#	0.00
16	Acenaphthylene	2.121	2.109	0.6	180#	-0.02
17	Acenaphthene	1.368	1.336	2.3	174#	-0.02
18	Fluorene	1.753	1.717	2.1	174#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	155#	-0.01
20	4-Bromophenyl-phenylether	0.243	0.258	-6.2	167#	-0.01
21	Hexachlorobenzene	0.265	0.292	-10.2	172#	-0.01
22	Pentachlorophenol	0.098	0.105	-7.1	191#	-0.01
23	Phenanthrene	1.246	1.322	-6.1	166#	-0.01
24	Anthracene	1.125	1.206	-7.2	172#	0.00
25 SURR	Fluoranthene-d10	1.348	1.378	-2.2	160#	-0.01
26	Fluoranthene	1.549	1.598	-3.2	161#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	149	0.00
28	Pvrene	1.364	1.434	-5.1	162#	-0.01
29 S	Terphenyl-d14	0.983	1.041	-5.9	163#	-0.01
30	Benzo(a)anthracene	1.473	1.514	-2.8	159#	0.00
31	Chrysene	1.436	1.497	-4.2	159#	0.00
32	Bis(2-ethylhexyl)phthalate	0.758	0.698	7.9	153#	-0.02
33	Indeno(1,2,3-cd)pyrene	1.797	1.779	1.0	151#	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	145	-0.02
35	Benzo(b)fluoranthene	1.389	1.429	-2.9	158#	-0.02
36	Benzo(k)fluoranthene	1.373	1.405	-2.3	154#	-0.02
37 C	Benzo(a)pyrene	1.305	1.320	-1.1	154#	-0.02
38	Dibenzo(a,h)anthracene	1.335	1.315	1.5	149	-0.02
39	Benzo(a,h,i)perylene	1.321	1.318	0.2	150#	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN093019\
Data File : BN007890.D
Acq On : 30 Sep 2019 11:06
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 01 07:01:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			