

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN093019\
 Data File : BN007907.D
 Acq On : 30 Sep 2019 21:24
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 30 23:26:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	146	0.00
2	1,4-Dioxane	0.400	0.412	-3.0	178	0.00
3	n-Nitrosodimethylamine	0.400	0.541	-35.3#	130	0.00
4 S	2-Fluorophenol	0.400	0.366	8.5	154	0.00
5 S	Phenol-d6	0.400	0.390	2.5	158	0.00
6	bis(2-Chloroethyl)ether	0.400	0.489	-22.2	203	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	147	-0.01
8 S	Nitrobenzene-d5	0.400	0.426	-6.5	161	0.00
9	Naphthalene	0.400	0.426	-6.5	157	-0.01
10	Hexachlorobutadiene	0.400	0.435	-8.7	159	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.409	-2.2	155	-0.01
12	2-Methylnaphthalene	0.400	0.411	-2.7	155	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	129	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.329	17.8	126	-0.01
15 S	2-Fluorobiphenyl	0.400	0.414	-3.5	144	-0.01
16	Acenaphthylene	0.400	0.410	-2.5	147	-0.01
17	Acenaphthene	0.400	0.385	3.8	135	-0.01
18	Fluorene	0.400	0.377	5.8	132	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	-0.01
20	4-Bromophenyl-phenylether	0.400	0.445	-11.2	121	-0.02
21	Hexachlorobenzene	0.400	0.459	-14.8	124	-0.01
22	Pentachlorophenol	0.400	0.434	-8.5	134	-0.01
23	Phenanthrene	0.400	0.422	-5.5	114	-0.01
24	Anthracene	0.400	0.433	-8.2	120	-0.01
25 SURR	Fluoranthene-d10	0.400	0.421	-5.2	114	-0.01
26	Fluoranthene	0.400	0.420	-5.0	114	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	103	-0.01
28	Pvrene	0.400	0.442	-10.5	118	-0.01
29 S	Terphenyl-d14	0.400	0.452	-13.0	120	-0.01
30	Benzo(a)anthracene	0.400	0.430	-7.5	115	-0.01
31	Chrysene	0.400	0.414	-3.5	109	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.495	-23.7	142	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.401	-0.3	106	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	101	-0.02
35	Benzo(b)fluoranthene	0.400	0.405	-1.3	109	-0.02
36	Benzo(k)fluoranthene	0.400	0.414	-3.5	109	-0.02
37 C	Benzo(a)pyrene	0.400	0.409	-2.2	109	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.394	1.5	104	-0.03
39	Benzo(a,h,i)perylene	0.400	0.403	-0.8	106	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN093019\
Data File : BN007907.D
Acq On : 30 Sep 2019 21:24
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 30 23:26:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				