

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN093019\
 Data File : BN007907.D
 Acq On : 30 Sep 2019 21:24
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 30 23:26:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	146	0.00
2	1,4-Dioxane	0.446	0.459	-2.9	178#	0.00
3	n-Nitrosodimethylamine	0.204	0.276	-35.3#	130	0.00
4 S	2-Fluorophenol	1.377	1.261	8.4	154#	0.00
5 S	Phenol-d6	1.962	1.675	14.6	158#	0.00
6	bis(2-Chloroethyl)ether	1.133	1.385	-22.2	203#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	147	-0.01
8 S	Nitrobenzene-d5	0.350	0.373	-6.6	161#	0.00
9	Naphthalene	1.123	1.196	-6.5	157#	-0.01
10	Hexachlorobutadiene	0.236	0.257	-8.9	159#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.674	0.690	-2.4	155#	-0.01
12	2-Methylnaphthalene	0.759	0.779	-2.6	155#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	129	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.179	17.9	126	-0.01
15 S	2-Fluorobiphenyl	1.647	1.705	-3.5	144	-0.01
16	Acenaphthylene	2.121	2.174	-2.5	147	-0.01
17	Acenaphthene	1.368	1.315	3.9	135	-0.01
18	Fluorene	1.753	1.653	5.7	132	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	-0.01
20	4-Bromophenyl-phenylether	0.243	0.270	-11.1	121	-0.02
21	Hexachlorobenzene	0.265	0.304	-14.7	124	-0.01
22	Pentachlorophenol	0.098	0.106	-8.2	134	-0.01
23	Phenanthrene	1.246	1.314	-5.5	114	-0.01
24	Anthracene	1.125	1.218	-8.3	120	-0.01
25 SURR	Fluoranthene-d10	1.348	1.418	-5.2	114	-0.01
26	Fluoranthene	1.549	1.626	-5.0	114	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	103	-0.01
28	Pvrene	1.364	1.506	-10.4	118	-0.01
29 S	Terphenyl-d14	0.983	1.110	-12.9	120	-0.01
30	Benzo(a)anthracene	1.473	1.584	-7.5	115	-0.01
31	Chrysene	1.436	1.485	-3.4	109	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	0.937	-23.6	142	-0.02
33	Indeno(1,2,3-cd)pyrene	1.797	1.801	-0.2	106	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	101	-0.02
35	Benzo(b)fluoranthene	1.389	1.405	-1.2	109	-0.02
36	Benzo(k)fluoranthene	1.373	1.422	-3.6	109	-0.02
37 C	Benzo(a)pyrene	1.305	1.334	-2.2	109	-0.02
38	Dibenzo(a,h)anthracene	1.335	1.314	1.6	104	-0.03
39	Benzo(a,h,i)perylene	1.321	1.331	-0.8	106	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN093019\
Data File : BN007907.D
Acq On : 30 Sep 2019 21:24
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 30 23:26:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			