

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN093021\
 Data File : BN016584.D
 Acq On : 29 Sep 2021 20:55
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Sep 30 01:38:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN093021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 30 01:35:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	29447	0.400	ng	0.00
7) Naphthalene-d8	10.419	136	125607	0.400	ng	0.00
13) Acenaphthene-d10	14.281	164	64187	0.400	ng	0.00
19) Phenanthrene-d10	17.030	188	117029	0.400	ng	0.00
29) Chrysene-d12	21.238	240	224783	0.400	ng	0.00
35) Perylene-d12	23.498	264	227503	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.254	112	130747	1.785	ng	0.00
5) Phenol-d6	6.831	99	142992	1.943	ng	0.00
8) Nitrobenzene-d5	8.784	82	140366	2.105	ng	0.00
11) 2-Methylnaphthalene-d10	12.012	152	316929	1.774	ng	0.00
14) 2,4,6-Tribromophenol	15.779	330	54104	1.931	ng	0.00
15) 2-Fluorobiphenyl	12.899	172	438239	1.659	ng	0.00
27) Fluoranthene-d10	19.073	212	556879	1.761	ng	0.00
31) Terphenyl-d14	19.688	244	443693	0.916	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.263	88	23513	1.854	ng	# 75
3) n-Nitrosodimethylamine	3.576	42	40189	1.719	ng	97
6) bis(2-Chloroethyl)ether	7.089	93	139054	1.793	ng	100
9) Naphthalene	10.457	128	581598	1.716	ng	99
10) Hexachlorobutadiene	10.749	225	109765	1.641	ng	# 100
12) 2-Methylnaphthalene	12.083	142	368040	1.745	ng	99
16) Acenaphthylene	13.990	152	519875	1.827	ng	100
17) Acenaphthene	14.345	154	325954	1.686	ng	98
18) Fluorene	15.335	166	394581	1.675	ng	100
20) 4,6-Dinitro-2-methylph...	15.411	198	30605	4.289	ng	# 88
21) 4-Bromophenyl-phenylether	16.227	248	126173	1.823	ng	# 68
22) Hexachlorobenzene	16.336	284	141944	1.718	ng	100
23) Atrazine	16.506	200	100952	2.143	ng	99
24) Pentachlorophenol	16.689	266	56971	2.167	ng	100
25) Phenanthrene	17.066	178	603555	1.714	ng	100
26) Anthracene	17.164	178	546086	1.785	ng	100
28) Fluoranthene	19.103	202	677169	1.760	ng	100
30) Pyrene	19.470	202	693727	0.905	ng	100
32) Benzo(a)anthracene	21.227	228	730770	1.049	ng	99
33) Chrysene	21.281	228	732189	1.030	ng	99
34) Bis(2-ethylhexyl)phtha...	21.174	149	316429	1.525	ng	100
36) Indeno(1,2,3-cd)pyrene	25.778	276	905765	1.249	ng	99
37) Benzo(b)fluoranthene	22.817	252	762978	1.004	ng	100
38) Benzo(k)fluoranthene	22.862	252	750893	1.015	ng	# 96
39) Benzo(a)pyrene	23.395	252	703760	1.010	ng	99
40) Dibenzo(a,h)anthracene	25.802	278	735606	1.459	ng	98
41) Benzo(g,h,i)perylene	26.480	276	792099	1.201	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN093021\
Data File : BN016584.D
Acq On : 29 Sep 2021 20:55
Operator : CG/JU
Sample : SSTDIC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC1.6

Quant Time: Sep 30 01:38:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN093021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 01:35:49 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.652 min Scan# 496
 Delta R.T. -0.000 min
 Lab File: BN016584.D
 Acq: 29 Sep 2021 20:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 29447
 Ion Ratio Lower Upper
 152 100
 150 155.4 123.3 184.9
 115 57.5 46.2 69.2



#2
 1,4-Dioxane
 Concen: 1.854 ng
 RT: 3.263 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: BN016584.D
 Acq: 29 Sep 2021 20:55

Tgt Ion: 88 Resp: 23513
 Ion Ratio Lower Upper
 88 100
 43 31.8 49.1 73.7#
 58 74.4 52.0 78.0









#3
n-Nitrosodimethylamine
Concen: 1.719 ng
RT: 3.576 min Scan# 54
Delta R.T. -0.009 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 40189
Ion Ratio Lower Upper
42 100
74 151.2 123.7 185.5
44 6.9 6.3 9.5



#4
2-Fluorophenol
Concen: 1.785 ng
RT: 5.254 min Scan# 236
Delta R.T. -0.009 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion: 112 Resp: 130747
Ion Ratio Lower Upper
112 100
64 54.1 43.4 65.2
63 27.2 21.8 32.8









#5
Phenol-d6
Concen: 1.943 ng
RT: 6.831 min Scan# 407
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	99	Resp:	142992
Ion	Ratio	Lower	Upper
99	100		
42	16.6	13.3	19.9
71	29.5	23.7	35.5



#6
bis(2-Chloroethyl)ether
Concen: 1.793 ng
RT: 7.089 min Scan# 435
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:	93	Resp:	139054
Ion	Ratio	Lower	Upper
93	100		
63	73.4	59.1	88.7
95	32.7	26.3	39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.419 min Scan# 734
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	125607
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.8	13.2
54	4.9	4.0	6.0
68	4.3	3.6	5.4



#8
Nitrobenzene-d5
Concen: 2.105 ng
RT: 8.784 min Scan# 605
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:	82	Resp:	140366
Ion Ratio	Lower	Upper	
82	100		
128	37.4	28.6	42.8
54	51.0	42.0	63.0









#9
Naphthalene
Concen: 1.716 ng
RT: 10.457 min Scan# 737
Delta R.T. -0.013 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 581598
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.2 10.1 15.1



#10
Hexachlorobutadiene
Concen: 1.641 ng
RT: 10.749 min Scan# 760
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:225 Resp: 109765
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 1.774 ng
RT: 12.012 min Scan# 933
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 316929
Ion Ratio Lower Upper
152 100
151 20.2 16.3 24.5



#12
2-Methylnaphthalene
Concen: 1.745 ng
RT: 12.083 min Scan# 949
Delta R.T. -0.004 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:142 Resp: 368040
Ion Ratio Lower Upper
142 100
141 88.5 70.2 105.4
115 32.8 26.1 39.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.281 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 64187
Ion Ratio Lower Upper
164 100
162 99.9 79.8 119.6
160 44.4 35.7 53.5



#14
2,4,6-Tribromophenol
Concen: 1.931 ng
RT: 15.779 min Scan# 1302
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:330 Resp: 54104
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 31.7 25.4 38.2









#15
2-Fluorobiphenyl
Concen: 1.659 ng
RT: 12.899 min Scan# 1075
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	172	Resp:	438239
Ion	Ratio	Lower	Upper
172	100		
171	34.3	27.6	41.4
170	22.9	18.7	28.1



#16
Acenaphthylene
Concen: 1.827 ng
RT: 13.990 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:	152	Resp:	519875
Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.4	23.2
153	13.0	10.4	15.6









#17
Acenaphthene
Concen: 1.686 ng
RT: 14.345 min Scan# 1189
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 325954
Ion Ratio Lower Upper
154 100
153 111.7 90.5 135.7
152 53.9 44.5 66.7



#18
Fluorene
Concen: 1.675 ng
RT: 15.335 min Scan# 1267
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:166 Resp: 394581
Ion Ratio Lower Upper
166 100
165 97.2 78.1 117.1
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.030 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 117029
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.8 8.1 12.1



#20
4,6-Dinitro-2-methylphenol
Concen: 4.289 ng
RT: 15.411 min Scan# 1273
Delta R.T. -0.013 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:198 Resp: 30605
Ion Ratio Lower Upper
198 100
182 10.3 12.3 18.5#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.823 ng
RT: 16.227 min Scan# 1338
Delta R.T. -0.012 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 126173
Ion Ratio Lower Upper
248 100
250 90.7 83.7 125.5
141 82.9 25.3 37.9#



#22
Hexachlorobenzene
Concen: 1.718 ng
RT: 16.336 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:284 Resp: 141944
Ion Ratio Lower Upper
284 100
142 33.3 26.7 40.1
249 29.3 23.4 35.2









#23
Atrazine
Concen: 2.143 ng
RT: 16.506 min Scan# 1361
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 100952
Ion Ratio Lower Upper
200 100
173 30.4 25.0 37.6
215 46.2 37.1 55.7



#24
Pentachlorophenol
Concen: 2.167 ng
RT: 16.689 min Scan# 1376
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:266 Resp: 56971
Ion Ratio Lower Upper
266 100
264 62.7 50.5 75.7
268 63.9 51.4 77.2









#25
Phenanthrene
Concen: 1.714 ng
RT: 17.066 min Scan# 1407
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 603555
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.6
179 15.3 12.2 18.2



#26
Anthracene
Concen: 1.785 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:178 Resp: 546086
Ion Ratio Lower Upper
178 100
176 18.2 14.6 22.0
179 15.4 12.3 18.5









#27
Fluoranthene-d10
Concen: 1.761 ng
RT: 19.073 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 556879
Ion Ratio Lower Upper
212 100
106 13.8 11.0 16.6
104 7.5 6.0 9.0



#28
Fluoranthene
Concen: 1.760 ng
RT: 19.103 min Scan# 1704
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:202 Resp: 677169
Ion Ratio Lower Upper
202 100
101 11.5 9.1 13.7
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.238 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 224783
Ion Ratio Lower Upper
240 100
120 13.4 11.6 17.4
236 26.9 21.6 32.4



#30
Pyrene
Concen: 0.905 ng
RT: 19.470 min Scan# 1778
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:202 Resp: 693727
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.5 13.9 20.9









#31
Terphenyl-d14
Concen: 0.916 ng
RT: 19.688 min Scan# 1822
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 443693
Ion Ratio Lower Upper
244 100
212 7.1 5.8 8.8
122 11.1 9.2 13.8



#32
Benzo(a)anthracene
Concen: 1.049 ng
RT: 21.227 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:228 Resp: 730770
Ion Ratio Lower Upper
228 100
226 26.1 21.4 32.0
229 20.1 15.6 23.4









#33
Chrysene
Concen: 1.030 ng
RT: 21.281 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	228	Resp:	732189
Ion Ratio	Lower	Upper	
228	100		
226	28.9	23.4	35.0
229	19.9	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.525 ng
RT: 21.174 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:	149	Resp:	316429
Ion Ratio	Lower	Upper	
149	100		
167	27.6	22.0	33.0
279	4.0	3.1	4.7









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.498 min Scan# 2431
 Delta R.T. 0.000 min
 Lab File: BN016584.D
 Acq: 29 Sep 2021 20:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:264 Resp: 227503
 Ion Ratio Lower Upper
 264 100
 260 24.6 19.7 29.5
 265 22.5 18.9 28.3



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 1.249 ng
 RT: 25.778 min Scan# 3097
 Delta R.T. -0.007 min
 Lab File: BN016584.D
 Acq: 29 Sep 2021 20:55

Tgt Ion:276 Resp: 905765
 Ion Ratio Lower Upper
 276 100
 138 25.4 21.1 31.7
 277 25.0 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 1.004 ng
RT: 22.817 min Scan# 2232
Delta R.T. -0.003 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 762978
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 10.2 8.2 12.2



#38
Benzo(k)fluoranthene
Concen: 1.015 ng
RT: 22.862 min Scan# 2245
Delta R.T. -0.003 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Tgt Ion:252 Resp: 750893
Ion Ratio Lower Upper
252 100
253 21.8 18.2 27.2
125 10.9 11.0 16.4#





#39

Benzo(a)pyrene

Concen: 1.010 ng

RT: 23.395 min Scan# 2401

Delta R.T. -0.003 min

Lab File: BN016584.D

Acq: 29 Sep 2021 20:55





Tgt Ion:252 Resp: 703760

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.6

125 11.1 9.2 13.8



#40

Dibenzo(a,h)anthracene

Concen: 1.459 ng

RT: 25.802 min Scan# 3104

Delta R.T. -0.003 min

Lab File: BN016584.D

Acq: 29 Sep 2021 20:55




Tgt Ion:278 Resp: 735606


Ion Ratio Lower Upper

278 100

139 17.1 15.6 23.4

279 23.7 19.0 28.4





#41
Benzo(g,h,i)perylene
Concen: 1.201 ng
RT: 26.480 min Scan# 3302
Delta R.T. 0.000 min
Lab File: BN016584.D
Acq: 29 Sep 2021 20:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:276 Resp: 792099
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.6
138 21.7 17.5 26.3

