

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN100118\  
Data File : BN002877.D  
Acq On : 01 Oct 2018 13:52  
Operator : MJ/SJ  
Sample : J5117-03  
Misc :  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
JLC79

Manual Integrations  
APPROVED

Sohil  
10/2/2018 5:45:39 PM

Quant Time: Oct 01 14:59:58 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN092018.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Oct 01 14:40:21 2018  
Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.64  | 152  | 1706     | 0.40 | ng/ul | -0.02    |
| 2) Naphthalene-d8         | 10.40 | 136  | 7026     | 0.40 | ng/ul | -0.02    |
| 6) Acenaphthene-d10       | 14.27 | 164  | 4284     | 0.40 | ng/ul | 0.00     |
| 10) Phenanthrene-d10      | 17.02 | 188  | 9193m    | 0.40 | ng/ul | -0.01    |
| 16) Chrysene-d12          | 21.24 | 240  | 7844m    | 0.40 | ng/ul | 0.00     |
| 20) Perylene-d12          | 23.44 | 264  | 7720     | 0.40 | ng/ul | -0.01    |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|------|-------|----------|
| 4) 2-Methylnaphthalene-d10  | 12.03 | 152  | 4495     | 0.44 | ng/ul | 0.00     |
| 14) Fluoranthene-d10        | 19.06 | 212  | 11065m   | 0.47 | ng/ul | -0.01    |

| Target Compounds | Ovalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN100118\  
Data File : BN002877.D  
Acq On : 01 Oct 2018 13:52  
Operator : MJ/SJ  
Sample : J5117-03  
Misc :  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
JLC79

Quant Time: Oct 01 14:59:58 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN092018.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Oct 01 14:40:21 2018  
Response via : Initial Calibration

Manual Integrations  
APPROVED

Sohil  
10/2/2018 5:45:39 PM

