

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100125\
 Data File : BN037981.D
 Acq On : 01 Oct 2025 16:39
 Operator : RC/JU
 Sample : SSTD3.290
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2223

Quant Time: Oct 02 09:34:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 02 09:28:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.808	152	7791	0.400	ng/ul	0.00
4) Naphthalene-d8	10.610	136	24375	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.468	164	13227	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.212	188	24823	0.400	ng/ul	0.00
17) Chrysene-d12	21.395	240	21945	0.400	ng/ul	0.00
23) Perylene-d12	23.716	264	16490	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	25920	3.554	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.285	88	27603	2.986	ng/ul#	88
14) Pentachlorophenol	16.866	266	21811	2.507	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100125\
Data File : BN037981.D
Acq On : 01 Oct 2025 16:39
Operator : RC/JU
Sample : SSTD3.220
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2223

Quant Time: Oct 02 09:34:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 02 09:28:18 2025
Response via : Initial Calibration

