

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100223\
 Data File : BN028003.D
 Acq On : 03 Oct 2023 00:50
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 03 01:31:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 02 19:18:04 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	102	0.00
2	1,4-Dioxane	0.400	0.450	-12.5	112	0.00
3	n-Nitrosodimethylamine	0.400	0.397	0.8	101	0.00
4 S	2-Fluorophenol	0.400	0.367	8.3	100	0.00
5 S	Phenol-d6	0.400	0.360	10.0	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.395	1.3	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.371	7.3	98	0.00
9	Naphthalene	0.400	0.398	0.5	101	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.390	2.5	99	0.00
12	2-Methylnaphthalene	0.400	0.398	0.5	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.350	12.5	94	0.00
15 S	2-Fluorobiphenyl	0.400	0.388	3.0	99	0.00
16	Acenaphthylene	0.400	0.386	3.5	97	0.00
17	Acenaphthene	0.400	0.404	-1.0	100	0.00
18	Fluorene	0.400	0.408	-2.0	100	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.318	20.5	82	0.00
21	4-Bromophenyl-phenylether	0.400	0.394	1.5	102	0.00
22	Hexachlorobenzene	0.400	0.407	-1.7	102	0.00
23	Atrazine	0.400	0.385	3.8	99	0.00
24	Pentachlorophenol	0.400	0.298	25.5#	90	0.00
25	Phenanthrene	0.400	0.406	-1.5	104	-0.01
26	Anthracene	0.400	0.392	2.0	100	0.00
27 SURR	Fluoranthene-d10	0.400	0.428	-7.0	105	0.00
28	Fluoranthene	0.400	0.431	-7.7	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	113	0.00
30	Pyrene	0.400	0.367	8.3	106	0.00
31 S	Terphenyl-d14	0.400	0.355	11.3	107	0.00
32	Benzo(a)anthracene	0.400	0.402	-0.5	110	0.00
33	Chrysene	0.400	0.416	-4.0	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.363	9.3	100	0.00
35 I	Perylene-d12	0.400	0.400	0.0	135	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.395	1.3	145	-0.01
37	Benzo(b)fluoranthene	0.400	0.384	4.0	129	0.00
38	Benzo(k)fluoranthene	0.400	0.388	3.0	130	-0.01
39 C	Benzo(a)pyrene	0.400	0.394	1.5	133	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.400	0.0	147	-0.01
41	Benzo(g,h,i)perylene	0.400	0.383	4.3	143	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4