

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100322\
 Data File : BN021980.D
 Acq On : 03 Oct 2022 16:14
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN100322

Quant Time: Oct 04 01:36:00 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 04 01:32:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.382	4.5	97	0.00
3	n-Nitrosodimethylamine	0.400	0.361	9.8	95	0.00
4 S	2-Fluorophenol	0.400	0.371	7.3	98	0.00
5 S	Phenol-d6	0.400	0.363	9.3	98	0.00
6	bis(2-Chloroethyl)ether	0.400	0.373	6.8	99	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.357	10.8	98	0.00
9	Naphthalene	0.400	0.372	7.0	101	0.00
10	Hexachlorobutadiene	0.400	0.372	7.0	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.374	6.5	102	0.00
12	2-Methylnaphthalene	0.400	0.352	12.0	106	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.400	0.353	11.8	107	0.00
15 S	2-Fluorobiphenyl	0.400	0.379	5.3	102	0.00
16	Acenaphthylene	0.400	0.348	13.0	102	0.00
17	Acenaphthene	0.400	0.373	6.8	104	0.00
18	Fluorene	0.400	0.367	8.3	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.414	-3.5	107	0.00
21	4-Bromophenyl-phenylether	0.400	0.359	10.3	103	0.00
22	Hexachlorobenzene	0.400	0.372	7.0	106	0.00
23	Atrazine	0.400	0.351	12.3	108	0.00
24	Pentachlorophenol	0.400	0.330	17.5	109	0.00
25	Phenanthrene	0.400	0.367	8.3	105	0.00
26	Anthracene	0.400	0.346	13.5	104	0.00
27 SURR	Fluoranthene-d10	0.400	0.361	9.8	106	0.00
28	Fluoranthene	0.400	0.361	9.8	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.379	5.3	105	0.00
31 S	Terphenyl-d14	0.400	0.386	3.5	107	0.00
32	Benzo(a)anthracene	0.400	0.362	9.5	105	0.00
33	Chrysene	0.400	0.384	4.0	105	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.348	13.0	104	0.00
35 I	Perylene-d12	0.400	0.400	0.0	105	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.354	11.5	103	0.00
37	Benzo(b)fluoranthene	0.400	0.369	7.8	107	0.00
38	Benzo(k)fluoranthene	0.400	0.360	10.0	102	0.00
39 C	Benzo(a)pyrene	0.400	0.347	13.3	101	0.00
40	Dibenzo(a,h)anthracene	0.400	0.355	11.3	103	0.00
41	Benzo(g,h,i)perylene	0.400	0.364	9.0	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN100322