

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100322\
 Data File : BN021983.D
 Acq On : 03 Oct 2022 19:55
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 04 01:51:09 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 04 01:32:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.371	7.3	95	0.00
3	n-Nitrosodimethylamine	0.400	0.367	8.3	98	0.00
4 S	2-Fluorophenol	0.400	0.382	4.5	102	0.00
5 S	Phenol-d6	0.400	0.372	7.0	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.374	6.5	100	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	99	0.00
8 S	Nitrobenzene-d5	0.400	0.358	10.5	98	0.00
9	Naphthalene	0.400	0.372	7.0	100	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.358	10.5	97	0.00
12	2-Methylnaphthalene	0.400	0.355	11.3	106	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.400	0.377	5.8	112	0.00
15 S	2-Fluorobiphenyl	0.400	0.381	4.8	100	0.00
16	Acenaphthylene	0.400	0.344	14.0	98	0.00
17	Acenaphthene	0.400	0.367	8.3	99	0.00
18	Fluorene	0.400	0.379	5.3	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.424	-6.0	104	0.00
21	4-Bromophenyl-phenylether	0.400	0.373	6.8	100	0.00
22	Hexachlorobenzene	0.400	0.378	5.5	101	0.00
23	Atrazine	0.400	0.362	9.5	105	0.00
24	Pentachlorophenol	0.400	0.354	11.5	109	0.00
25	Phenanthrene	0.400	0.373	6.8	100	0.00
26	Anthracene	0.400	0.353	11.8	99	0.00
27 SURR	Fluoranthene-d10	0.400	0.366	8.5	101	0.00
28	Fluoranthene	0.400	0.368	8.0	101	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	102	0.00
30	Pyrene	0.400	0.375	6.3	101	0.00
31 S	Terphenyl-d14	0.400	0.375	6.3	101	0.00
32	Benzo(a)anthracene	0.400	0.361	9.8	101	0.00
33	Chrysene	0.400	0.372	7.0	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.361	9.8	104	0.00
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.359	10.3	109	0.00
37	Benzo(b)fluoranthene	0.400	0.352	12.0	107	0.00
38	Benzo(k)fluoranthene	0.400	0.348	13.0	103	0.00
39 C	Benzo(a)pyrene	0.400	0.356	11.0	109	0.00
40	Dibenzo(a,h)anthracene	0.400	0.354	11.5	107	0.00
41	Benzo(g,h,i)perylene	0.400	0.339	15.3	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4