

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100322\
 Data File : BN021999.D
 Acq On : 04 Oct 2022 05:37
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 04 06:12:39 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 04 01:32:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.518	0.499	3.7	104	0.00
3	n-Nitrosodimethylamine	0.563	0.539	4.3	108	0.00
4 S	2-Fluorophenol	0.927	0.961	-3.7	117	0.00
5 S	Phenol-d6	1.202	1.237	-2.9	118	0.00
6	bis(2-Chloroethyl)ether	1.271	1.249	1.7	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.322	0.300	6.8	112	0.00
9	Naphthalene	1.208	1.188	1.7	116	-0.01
10	Hexachlorobutadiene	0.209	0.197	5.7	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.735	0.703	4.4	114	0.00
12	2-Methylnaphthalene	0.179	0.192	-7.3	141	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.153	0.152	0.7	136	0.00
15 S	2-Fluorobiphenyl	1.748	1.652	5.5	115	-0.01
16	Acenaphthylene	1.903	1.654	13.1	115	0.00
17	Acenaphthene	1.338	1.246	6.9	117	0.00
18	Fluorene	1.613	1.543	4.3	123	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	-0.01
20	4,6-Dinitro-2-methylphenol	0.057	0.050	12.3	134	0.00
21	4-Bromophenyl-phenylether	0.244	0.225	7.8	116	0.00
22	Hexachlorobenzene	0.303	0.281	7.3	116	0.00
23	Atrazine	0.181	0.167	7.7	124	0.00
24	Pentachlorophenol	0.093	0.095	-2.2	148	-0.01
25	Phenanthrene	1.375	1.301	5.4	119	0.00
26	Anthracene	1.104	0.999	9.5	119	0.00
27 SURR	Fluoranthene-d10	1.063	0.996	6.3	121	0.00
28	Fluoranthene	1.473	1.369	7.1	119	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
30	Pyrene	1.757	1.673	4.8	119	0.00
31 S	Terphenyl-d14	0.804	0.779	3.1	121	0.00
32	Benzo(a)anthracene	1.504	1.409	6.3	122	0.00
33	Chrysene	1.669	1.560	6.5	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.722	0.676	6.4	126	0.00
35 I	Perylene-d12	1.000	1.000	0.0	121	0.00
36	Indeno(1,2,3-cd)pyrene	2.021	1.740	13.9	115	0.00
37	Benzo(b)fluoranthene	1.896	1.733	8.6	122	0.00
38	Benzo(k)fluoranthene	1.888	1.686	10.7	116	0.00
39 C	Benzo(a)pyrene	1.446	1.311	9.3	122	0.00
40	Dibenzo(a,h)anthracene	1.612	1.371	15.0	113	0.00
41	Benzo(g,h,i)perylene	1.740	1.441	17.2	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4