

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100323\
 Data File : BN028011.D
 Acq On : 03 Oct 2023 15:19
 Operator : MA/JU
 Sample : SSTD3.242
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2214

Quant Time: Oct 03 15:47:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 03 14:19:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.950	152	4133	0.400	ng/ul	0.00
4) Naphthalene-d8	10.768	136	14258	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.592	164	8549	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.346	188	18249	0.400	ng/ul	0.00
17) Chrysene-d12	21.536	240	17860	0.400	ng/ul	0.00
23) Perylene-d12	24.006	264	17812	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.321	96	16423	3.071	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.359	88	17166	3.074	ng/ul#	85
14) Pentachlorophenol	16.979	266	15159	3.520	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100323\
Data File : BN028011.D
Acq On : 03 Oct 2023 15:19
Operator : MA/JU
Sample : SSTD3.242
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2214

Quant Time: Oct 03 15:47:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 03 14:19:11 2023
Response via : Initial Calibration

