

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100323\
 Data File : BN028032.D
 Acq On : 04 Oct 2023 04:55
 Operator : MA/JU
 Sample : 04497-12
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BBJ14RE

Quant Time: Oct 04 05:22:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 03 15:50:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.945	152	5723	0.400	ng/ul	0.00
4) Naphthalene-d8	10.773	136	19864	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.606	164	193185	0.400	ng/ul	# 0.01
13) Phenanthrene-d10	17.350	188	25255	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.533	240	19288	0.400	ng/ul	# 0.00
23) Perylene-d12	24.003	264	17533	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.321	96	19398	2.699	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.346	152	10298	0.322	ng/ul	0.00
18) Fluoranthene-d10	19.376	212	18850	0.308	ng/ul	0.00
Target Compounds				Qvalue		
10) Acenaphthylene	14.333	152	2301258	2.626	ng/ul#	71
19) Fluoranthene	19.404	202	2469	0.034	ng/ul#	54
20) Pyrene	19.771	202	2514	0.033	ng/ul#	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100323\
Data File : BN028032.D
Acq On : 04 Oct 2023 04:55
Operator : MA/JU
Sample : 04497-12
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BBJ14RE

Quant Time: Oct 04 05:22:25 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 03 15:50:04 2023
Response via : Initial Calibration

