

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100419\
 Data File : BN008004.D
 Acq On : 04 Oct 2019 15:35
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 05 03:16:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 03 18:39:49 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	-0.02
2	1,4-Dioxane	0.446	0.464	-4.0	112	-0.03
3	n-Nitrosodimethylamine	0.204	0.329	-61.3#	96	-0.03
4 S	2-Fluorophenol	1.377	1.207	12.3	92	-0.02
5 S	Phenol-d6	1.962	1.531	22.0	90	0.00
6	bis(2-Chloroethyl)ether	1.133	1.215	-7.2	111	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	91	-0.03
8 S	Nitrobenzene-d5	0.350	0.344	1.7	92	-0.01
9	Naphthalene	1.123	1.149	-2.3	93	-0.03
10	Hexachlorobutadiene	0.236	0.262	-11.0	100	-0.04
11 SURR	2-Methylnaphthalene-d10	0.674	0.699	-3.7	97	-0.02
12	2-Methylnaphthalene	0.759	0.789	-4.0	97	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	-0.02
14 S	2,4,6-Tribromophenol	0.218	0.190	12.8	91	-0.02
15 S	2-Fluorobiphenyl	1.647	1.614	2.0	92	-0.02
16	Acenaphthylene	2.121	1.990	6.2	91	-0.03
17	Acenaphthene	1.368	1.259	8.0	87	-0.02
18	Fluorene	1.753	1.608	8.3	87	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	81	-0.02
20	4-Bromophenyl-phenylether	0.243	0.246	-1.2	83	-0.03
21	Hexachlorobenzene	0.265	0.287	-8.3	88	-0.02
22	Pentachlorophenol	0.098	0.085	13.3	81	-0.02
23	Phenanthrene	1.246	1.236	0.8	81	-0.02
24	Anthracene	1.125	1.136	-1.0	84	-0.02
25 SURR	Fluoranthene-d10	1.348	1.340	0.6	81	-0.02
26	Fluoranthene	1.549	1.529	1.3	81	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	74	-0.02
28	Pyrene	1.364	1.470	-7.8	83	-0.02
29 S	Terphenyl-d14	0.983	1.051	-6.9	82	-0.02
30	Benzo(a)anthracene	1.473	1.501	-1.9	79	-0.02
31	Chrysene	1.436	1.460	-1.7	77	-0.02
32	Bis(2-ethylhexyl)phthalate	0.758	0.841	-10.9	92	-0.03
33	Indeno(1,2,3-cd)pyrene	1.797	1.781	0.9	75	-0.05
34 I	Perylene-d12	1.000	1.000	0.0	72	-0.03
35	Benzo(b)fluoranthene	1.389	1.369	1.4	75	-0.03
36	Benzo(k)fluoranthene	1.373	1.424	-3.7	78	-0.03
37 C	Benzo(a)pyrene	1.305	1.324	-1.5	77	-0.03
38	Dibenzo(a,h)anthracene	1.335	1.328	0.5	75	-0.05
39	Benzo(g,h,i)perylene	1.321	1.328	-0.5	76	-0.05

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100419\
Data File : BN008004.D
Acq On : 04 Oct 2019 15:35
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 05 03:16:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 03 18:39:49 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			