

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100419\
 Data File : BN008021.D
 Acq On : 05 Oct 2019 01:44
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 05 04:31:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 03 18:39:49 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	-0.02
2	1,4-Dioxane	0.446	0.564	-26.5#	132	-0.03
3	n-Nitrosodimethylamine	0.204	0.329	-61.3#	93	-0.02
4 S	2-Fluorophenol	1.377	1.297	5.8	95	0.00
5 S	Phenol-d6	1.962	1.659	15.4	94	0.02
6	bis(2-Chloroethyl)ether	1.133	1.580	-39.5#	139	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	88	-0.01
8 S	Nitrobenzene-d5	0.350	0.355	-1.4	92	-0.01
9	Naphthalene	1.123	1.154	-2.8	91	-0.01
10	Hexachlorobutadiene	0.236	0.271	-14.8	100	-0.03
11 SURR	2-Methylnaphthalene-d10	0.674	0.696	-3.3	94	-0.01
12	2-Methylnaphthalene	0.759	0.798	-5.1	95	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	81	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.240	-10.1	107	-0.01
15 S	2-Fluorobiphenyl	1.647	1.614	2.0	86	-0.01
16	Acenaphthylene	2.121	3.095	-45.9#	132	-0.02
17	Acenaphthene	1.368	1.368	0.0	89	-0.02
18	Fluorene	1.753	1.954	-11.5	99	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	73	-0.01
20	4-Bromophenyl-phenylether	0.243	0.258	-6.2	79	-0.02
21	Hexachlorobenzene	0.265	0.285	-7.5	79	-0.01
22	Pentachlorophenol	0.098	0.125	-27.6#	108	0.00
23	Phenanthrene	1.246	1.315	-5.5	78	-0.01
24	Anthracene	1.125	1.251	-11.2	84	-0.01
25 SURR	Fluoranthene-d10	1.348	1.383	-2.6	76	0.00
26	Fluoranthene	1.549	1.656	-6.9	79	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	67	-0.01
28	Pyrene	1.364	1.598	-17.2	81	-0.01
29 S	Terphenyl-d14	0.983	1.087	-10.6	76	-0.01
30	Benzo(a)anthracene	1.473	1.569	-6.5	74	-0.01
31	Chrysene	1.436	1.448	-0.8	69	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	1.049	-38.4#	103	-0.03
33	Indeno(1,2,3-cd)pyrene	1.797	1.767	1.7	67	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	66	-0.02
35	Benzo(b)fluoranthene	1.389	1.385	0.3	70	-0.02
36	Benzo(k)fluoranthene	1.373	1.321	3.8	66	-0.02
37 C	Benzo(a)pyrene	1.305	1.346	-3.1	72	-0.02
38	Dibenzo(a,h)anthracene	1.335	1.303	2.4	68	-0.02
39	Benzo(g,h,i)perylene	1.321	1.316	0.4	69	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100419\
Data File : BN008021.D
Acq On : 05 Oct 2019 01:44
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 05 04:31:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 03 18:39:49 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			