

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100423\
 Data File : BN028036.D
 Acq On : 04 Oct 2023 15:03
 Operator : MA/JU
 Sample : PB155913BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK913

Quant Time: Oct 05 02:05:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 03 15:50:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.941	152	4445	0.400	ng/ul	0.00
4) Naphthalene-d8	10.762	136	14817	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.587	164	9295	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.342	188	18770	0.400	ng/ul	0.00
17) Chrysene-d12	21.530	240	12134	0.400	ng/ul #	0.00
23) Perylene-d12	23.997	264	13183	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.321	96	42876	7.680	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.341	152	10050	0.421	ng/ul	-0.01
18) Fluoranthene-d10	19.371	212	18905	0.491	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100423\
Data File : BN028036.D
Acq On : 04 Oct 2023 15:03
Operator : MA/JU
Sample : PB155913BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK913

Quant Time: Oct 05 02:05:03 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 03 15:50:04 2023
Response via : Initial Calibration

