

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100519\
 Data File : BN008023.D
 Acq On : 05 Oct 2019 13:29
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 05:32:39 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 03 18:39:49 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	83	-0.02
2	1,4-Dioxane	0.400	0.434	-8.5	106	0.00
3	n-Nitrosodimethylamine	0.400	0.522	-30.5#	71	-0.03
4 S	2-Fluorophenol	0.400	0.377	5.8	90	0.00
5 S	Phenol-d6	0.400	0.387	3.3	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.509	-27.2#	120	-0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	82	-0.01
8 S	Nitrobenzene-d5	0.400	0.453	-13.2	95	-0.01
9	Naphthalene	0.400	0.412	-3.0	85	-0.03
10	Hexachlorobutadiene	0.400	0.456	-14.0	92	-0.04
11 SURR	2-Methylnaphthalene-d10	0.400	0.421	-5.2	89	-0.02
12	2-Methylnaphthalene	0.400	0.419	-4.7	88	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.419	-4.7	105	-0.01
15 S	2-Fluorobiphenyl	0.400	0.364	9.0	82	-0.02
16	Acenaphthylene	0.400	0.364	9.0	85	-0.02
17	Acenaphthene	0.400	0.352	12.0	81	-0.02
18	Fluorene	0.400	0.356	11.0	81	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	72	-0.01
20	4-Bromophenyl-phenylether	0.400	0.427	-6.7	78	-0.02
21	Hexachlorobenzene	0.400	0.434	-8.5	79	-0.02
22	Pentachlorophenol	0.400	0.427	-6.7	88	0.00
23	Phenanthrene	0.400	0.408	-2.0	75	-0.02
24	Anthracene	0.400	0.435	-8.7	81	-0.01
25 SURR	Fluoranthene-d10	0.400	0.408	-2.0	74	-0.02
26	Fluoranthene	0.400	0.409	-2.2	75	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	63	-0.02
28	Pyrene	0.400	0.465	-16.3	76	-0.02
29 S	Terphenyl-d14	0.400	0.456	-14.0	75	-0.02
30	Benzo(a)anthracene	0.400	0.418	-4.5	69	-0.01
31	Chrysene	0.400	0.417	-4.2	67	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.521	-30.3#	92	-0.03
33	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	61	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	60	-0.02
35	Benzo(b)fluoranthene	0.400	0.398	0.5	63	-0.02
36	Benzo(k)fluoranthene	0.400	0.405	-1.3	63	-0.02
37 C	Benzo(a)pyrene	0.400	0.401	-0.3	63	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.391	2.3	61	-0.03
39	Benzo(g,h,i)perylene	0.400	0.393	1.8	61	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100519\
Data File : BN008023.D
Acq On : 05 Oct 2019 13:29
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 07 05:32:39 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 03 18:39:49 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				