

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100519\
 Data File : BN008042.D
 Acq On : 06 Oct 2019 02:37
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 07:33:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 03 18:39:49 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	-0.02
2	1,4-Dioxane	0.446	0.522	-17.0	89	-0.03
3	n-Nitrosodimethylamine	0.204	0.326	-59.8#	67	-0.03
4 S	2-Fluorophenol	1.377	1.268	7.9	68	-0.02
5 S	Phenol-d6	1.962	1.560	20.5	64	0.00
6	bis(2-Chloroethyl)ether	1.133	1.387	-22.4	89	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	65	-0.03
8 S	Nitrobenzene-d5	0.350	0.350	0.0	67	-0.01
9	Naphthalene	1.123	1.145	-2.0	66	-0.03
10	Hexachlorobutadiene	0.236	0.271	-14.8	74	-0.04
11 SURR	2-Methylnaphthalene-d10	0.674	0.704	-4.5	70	-0.02
12	2-Methylnaphthalene	0.759	0.791	-4.2	69	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	64	-0.02
14 S	2,4,6-Tribromophenol	0.218	0.213	2.3	74	-0.01
15 S	2-Fluorobiphenyl	1.647	1.515	8.0	63	-0.02
16	Acenaphthylene	2.121	2.106	0.7	70	-0.02
17	Acenaphthene	1.368	1.241	9.3	63	-0.02
18	Fluorene	1.753	1.610	8.2	64	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	57	-0.02
20	4-Bromophenyl-phenylether	0.243	0.245	-0.8	59	-0.02
21	Hexachlorobenzene	0.265	0.294	-10.9	64	-0.02
22	Pentachlorophenol	0.098	0.100	-2.0	67	0.00
23	Phenanthrene	1.246	1.217	2.3	57	-0.02
24	Anthracene	1.125	1.143	-1.6	61	-0.01
25 SURR	Fluoranthene-d10	1.348	1.347	0.1	58	-0.02
26	Fluoranthene	1.549	1.564	-1.0	59	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	53	-0.02
28	Pyrene	1.364	1.508	-10.6	61	-0.02
29 S	Terphenyl-d14	0.983	1.041	-5.9	58	-0.02
30	Benzo(a)anthracene	1.473	1.476	-0.2	55	-0.02
31	Chrysene	1.436	1.463	-1.9	55	-0.02
32	Bis(2-ethylhexyl)phthalate	0.758	0.926	-22.2	72	-0.03
33	Indeno(1,2,3-cd)pyrene	1.797	1.781	0.9	54	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	52	-0.03
35	Benzo(b)fluoranthene	1.389	1.395	-0.4	55	-0.03
36	Benzo(k)fluoranthene	1.373	1.312	4.4	52	-0.03
37 C	Benzo(a)pyrene	1.305	1.303	0.2	55	-0.03
38	Dibenzo(a,h)anthracene	1.335	1.316	1.4	54	-0.04
39	Benzo(g,h,i)perylene	1.321	1.341	-1.5	55	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100519\
Data File : BN008042.D
Acq On : 06 Oct 2019 02:37
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 07 07:33:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 03 18:39:49 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			