

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100519\
 Data File : BN008048.D
 Acq On : 06 Oct 2019 06:48
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 07:33:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 03 18:39:49 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	65	-0.02
2	1,4-Dioxane	0.400	0.480	-20.0	92	-0.03
3	n-Nitrosodimethylamine	0.400	0.655	-63.7#	70	-0.03
4 S	2-Fluorophenol	0.400	0.364	9.0	68	-0.02
5 S	Phenol-d6	0.400	0.359	10.3	64	0.00
6	bis(2-Chloroethyl)ether	0.400	0.482	-20.5	89	-0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	66	-0.03
8 S	Nitrobenzene-d5	0.400	0.386	3.5	66	-0.01
9	Naphthalene	0.400	0.407	-1.7	68	-0.03
10	Hexachlorobutadiene	0.400	0.459	-14.8	76	-0.04
11 SURR	2-Methylnaphthalene-d10	0.400	0.419	-4.7	72	-0.02
12	2-Methylnaphthalene	0.400	0.415	-3.7	71	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	65	-0.03
14 S	2,4,6-Tribromophenol	0.400	0.397	0.8	77	-0.02
15 S	2-Fluorobiphenyl	0.400	0.365	8.8	64	-0.02
16	Acenaphthylene	0.400	0.401	-0.3	72	-0.03
17	Acenaphthene	0.400	0.367	8.3	65	-0.03
18	Fluorene	0.400	0.373	6.8	66	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	59	-0.01
20	4-Bromophenyl-phenylether	0.400	0.400	0.0	60	-0.03
21	Hexachlorobenzene	0.400	0.446	-11.5	66	-0.03
22	Pentachlorophenol	0.400	0.396	1.0	67	0.00
23	Phenanthrene	0.400	0.391	2.3	58	-0.03
24	Anthracene	0.400	0.410	-2.5	62	-0.01
25 SURR	Fluoranthene-d10	0.400	0.407	-1.7	61	-0.02
26	Fluoranthene	0.400	0.405	-1.3	60	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	54	-0.02
28	Pyrene	0.400	0.439	-9.7	62	-0.02
29 S	Terphenyl-d14	0.400	0.426	-6.5	60	-0.02
30	Benzo(a)anthracene	0.400	0.409	-2.2	58	-0.02
31	Chrysene	0.400	0.406	-1.5	57	-0.02
32	Bis(2-ethylhexyl)phthalate	0.400	0.487	-21.7	74	-0.04
33	Indeno(1,2,3-cd)pyrene	0.400	0.398	0.5	55	-0.04
34 I	Perylene-d12	0.400	0.400	0.0	54	-0.03
35	Benzo(b)fluoranthene	0.400	0.394	1.5	56	-0.03
36	Benzo(k)fluoranthene	0.400	0.390	2.5	55	-0.03
37 C	Benzo(a)pyrene	0.400	0.402	-0.5	57	-0.03
38	Dibenzo(a,h)anthracene	0.400	0.394	1.5	55	-0.04
39	Benzo(g,h,i)perylene	0.400	0.403	-0.8	57	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100519\
Data File : BN008048.D
Acq On : 06 Oct 2019 06:48
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 07 07:33:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 03 18:39:49 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				