

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100523\
 Data File : BN028112.D
 Acq On : 07 Oct 2023 09:19
 Operator : MA/JU
 Sample : PB156055BL
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Oct 08 00:50:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 06 05:15:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.933	152	4755	0.400	ng/ul	0.00
4) Naphthalene-d8	10.746	136	14535	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.578	164	7620	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.333	188	14980	0.400	ng/ul	0.00
17) Chrysene-d12	21.524	240	12470	0.400	ng/ul	0.00
23) Perylene-d12	23.986	264	14638	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.321	96	45138	7.558	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.335	152	7407	0.316	ng/ul	0.00
18) Fluoranthene-d10	19.362	212	13846	0.350	ng/ul	0.00
Target Compounds						
					Qvalue	
12) Fluorene	15.573	166	2215	0.072	ng/ul#	15
20) Pyrene	19.729	202	1968	0.040	ng/ul#	1
26) Benzo(a)pyrene	23.825	252	4479	0.097	ng/ul#	43


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100523\
Data File : BN028112.D
Acq On : 07 Oct 2023 09:19
Operator : MA/JU
Sample : PB156055BL
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :


Quant Time: Oct 08 00:50:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 06 05:15:40 2023
Response via : Initial Calibration







#12
Fluorene
Concen: 0.072 ng/ul
RT: 15.573 min Scan# 21
Delta R.T. -0.060 min
Lab File: BN028112.D
Acq: 07 Oct 2023 09:19

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:166 Resp: 2215
Ion Ratio Lower Upper
166 100
165 10.7 78.5 117.7#
167 37.9 11.2 16.8#



#20
Pyrene
Concen: 0.040 ng/ul
RT: 19.729 min Scan# 3727
Delta R.T. -0.028 min
Lab File: BN028112.D
Acq: 07 Oct 2023 09:19


Tgt Ion:202 Resp: 1968
Ion Ratio Lower Upper
202 100
101 162.0 11.2 16.8#
100 1152.4 9.1 13.7#





#26
Benzo(a)pyrene
Concen: 0.097 ng/ul
RT: 23.825 min Scan# 50
Delta R.T. -0.044 min
Lab File: BN028112.D
Acq: 07 Oct 2023 09:19

Instrument :
BNA_N
ClientSampleId :



Tgt Ion:252 Resp: 4479
Ion Ratio Lower Upper
252 100
253 57.9 23.4 35.2#
125 47.6 15.3 22.9#

