

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100523\
 Data File : BN028112.D
 Acq On : 07 Oct 2023 09:19
 Operator : MA/JU
 Sample : PB156055BL
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK055

Quant Time: Oct 08 00:50:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 06 05:15:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.933	152	4755	0.400	ng/ul	0.00
4) Naphthalene-d8	10.746	136	14535	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.578	164	7620	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.333	188	14980	0.400	ng/ul	0.00
17) Chrysene-d12	21.524	240	12470	0.400	ng/ul	0.00
23) Perylene-d12	23.986	264	14638	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.321	96	45138	7.558	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.335	152	7407	0.316	ng/ul	0.00
18) Fluoranthene-d10	19.362	212	13846	0.350	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100523\
Data File : BN028112.D
Acq On : 07 Oct 2023 09:19
Operator : MA/JU
Sample : PB156055BL
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK055

Quant Time: Oct 08 00:50:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 06 05:15:40 2023
Response via : Initial Calibration

