

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN100523\  
 Data File : BN028141.D  
 Acq On : 08 Oct 2023 03:31  
 Operator : MA/JU  
 Sample : 04624-01  
 Misc :  
 ALS Vial : 69 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 BBJQ0

Quant Time: Oct 08 03:57:55 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\SFAM-EPA-SIM-BN100323.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Oct 06 05:15:40 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|--------|------|----------|-------|--------|----------|
| -----                       |        |      |          |       |        |          |
| Internal Standards          |        |      |          |       |        |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.929  | 152  | 4310     | 0.400 | ng/ul  | 0.00     |
| 4) Naphthalene-d8           | 10.740 | 136  | 13236    | 0.400 | ng/ul  | -0.01    |
| 9) Acenaphthene-d10         | 14.569 | 164  | 7065     | 0.400 | ng/ul  | -0.01    |
| 13) Phenanthrene-d10        | 17.329 | 188  | 13930    | 0.400 | ng/ul  | 0.00     |
| 17) Chrysene-d12            | 21.518 | 240  | 11938    | 0.400 | ng/ul  | 0.00     |
| 23) Perylene-d12            | 23.977 | 264  | 13809    | 0.400 | ng/ul  | 0.00     |
| System Monitoring Compounds |        |      |          |       |        |          |
| 3) 1,4-Dioxane-d8           | 3.321  | 96   | 28848    | 5.329 | ng/ul  | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.330 | 152  | 7536     | 0.353 | ng/ul  | 0.00     |
| 18) Fluoranthene-d10        | 19.357 | 212  | 17073    | 0.451 | ng/ul  | 0.00     |
| Target Compounds            |        |      |          |       |        |          |
|                             |        |      |          |       | Qvalue |          |
| 2) 1,4-Dioxane              | 3.355  | 88   | 2214     | 0.386 | ng/ul  | 95       |
| -----                       |        |      |          |       |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN100523\  
Data File : BN028141.D  
Acq On : 08 Oct 2023 03:31  
Operator : MA/JU  
Sample : 04624-01  
Misc :  
ALS Vial : 69 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
BBJQ0

Quant Time: Oct 08 03:57:55 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\SFAM-EPA-SIM-BN100323.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Oct 06 05:15:40 2023  
Response via : Initial Calibration

