




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100523\
 Data File : BN028118.D
 Acq On : 07 Oct 2023 12:58
 Operator : MA/JU
 Sample : 04654-05
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHAU7

Manual IntegrationsAPPROVED

Quant Time: Oct 08 00:51:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 06 05:15:40 2023
 Response via : Initial Calibration

Reviewed By :Yogesh Patel 10/08/2023
 Supervised By :mohammad ahmed 10/08/2023



TIC: BN028118.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.330min (-3.330) 0.00 ng/u1

response 0

Ion	Exp%	Act%
96.00	100.00	0.00
64.00	62.70	0.00#
34.00	37.00	0.00#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100523\
 Data File : BN028118.D
 Acq On : 07 Oct 2023 12:58
 Operator : MA/JU
 Sample : 04654-05
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHAU7

Manual IntegrationsAPPROVED

Quant Time: Oct 08 00:51:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 06 05:15:40 2023
 Response via : Initial Calibration

Reviewed By :Yogesh Patel 10/08/2023
 Supervised By :mohammad ahmed 10/08/2023



TIC: BN028118.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.321min (-0.008) 5.13 ng/ul m

response 31681

Ion	Exp%	Act%
96.00	100.00	100.00
64.00	62.70	53.88
34.00	37.00	29.02#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100523\
 Data File : BN028118.D
 Acq On : 07 Oct 2023 12:58
 Operator : MA/JU
 Sample : 04654-05
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHAU7

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 10/08/2023
 Supervised By :mohammad ahmed 10/08/2023

Quant Time: Oct 08 00:51:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 06 05:15:40 2023
 Response via : Initial Calibration

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.933	152		4914	0.400	ng/ul	0.00
4) Naphthalene-d8	10.746	136		15021	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.573	164		8199	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.329	188		16304	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.522	240		13589	0.400	ng/ul	0.00
23) Perylene-d12	23.986	264		15600	0.400	ng/ul	0.00
System Monitoring Compounds							
3) 1,4-Dioxane-d8	3.321	96		31681m	5.133	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.330	152		7471	0.309	ng/ul	0.00
18) Fluoranthene-d10	19.357	212		17063	0.396	ng/ul	0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.359	88		4363	0.668	ng/ul	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual IntegrationsAPPROVED

Quant Time: Oct 08 00:51:22 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 06 05:15:40 2023
Response via : Initial Calibration

