

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100523\
 Data File : BN028066.D
 Acq On : 06 Oct 2023 02:26
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4289

Quant Time: Oct 06 05:15:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 06 05:15:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.937	152	6020	0.400	ng/ul	0.00
4) Naphthalene-d8	10.751	136	18378	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.583	164	9487	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.333	188	18585	0.400	ng/ul	0.00
17) Chrysene-d12	21.519	240	17050	0.400	ng/ul	0.00
23) Perylene-d12	23.977	264	19426	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.330	96	3118	0.412	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.335	152	11080	0.374	ng/ul	0.00
18) Fluoranthene-d10	19.362	212	21031	0.389	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.363	88	3183	0.398	ng/ul	96
5) Naphthalene	10.806	128	20653	0.396	ng/ul	100
7) 2-Methylnaphthalene	12.412	142	12344	0.375	ng/ul	99
8) 1-Methylnaphthalene	12.627	142	12701	0.373	ng/ul	99
10) Acenaphthylene	14.310	152	16993	0.395	ng/ul	100
11) Acenaphthene	14.643	153	13369	0.390	ng/ul	100
12) Fluorene	15.633	166	14503	0.381	ng/ul	99
14) Pentachlorophenol	16.970	266	1713	0.430	ng/ul	98
15) Phenanthrene	17.375	178	22004	0.397	ng/ul	99
16) Anthracene	17.468	178	20105	0.406	ng/ul	99
19) Fluoranthene	19.390	202	25296	0.389	ng/ul	99
20) Pyrene	19.757	202	26665	0.395	ng/ul	99
21) Benzo(a)anthracene	21.501	228	25313	0.407	ng/ul	99
22) Chrysene	21.557	228	25073	0.399	ng/ul	99
24) Benzo(b)fluoranthene	23.217	252	26715	0.371	ng/ul	93
25) Benzo(k)fluoranthene	23.267	252	27825	0.388	ng/ul#	93
26) Benzo(a)pyrene	23.869	252	24667	0.402	ng/ul#	91
27) Indeno(1,2,3-cd)pyrene	26.566	276	32340	0.435	ng/ul#	99
28) Dibenzo(a,h)anthracene	26.582	278	26228	0.434	ng/ul#	95
29) Benzo(g,h,i)perylene	27.377	276	26789	0.441	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100523\
Data File : BN028066.D
Acq On : 06 Oct 2023 02:26
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4289

Quant Time: Oct 06 05:15:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 06 05:15:40 2023
Response via : Initial Calibration

