


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100618\
Data File : BN002956.D
Acq On : 05 Oct 2018 07:59
Operator : MJ/SJ
Sample : DFTPP31
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP31

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091418MA.M
Title : SVOA CALIBRATION
Last Update : Mon Oct 08 14:32:35 2018



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2455


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.9	75474	PASS
68	69	0.00	2	1.8	1462	PASS
69	198	0.00	100	37.2	80458	PASS
70	69	0.00	2	0.6	445	PASS
127	198	10	80	45.2	97586	PASS
197	198	0.00	2	0.5	1000	PASS
198	198	100	100	100.0	216128	PASS
199	198	5	9	6.9	15009	PASS
275	198	10	60	26.9	58146	PASS
365	198	1	100	3.1	6687	PASS
441	198	0.01	100	13.1	28271	PASS
442	198	50	100	86.4	186712	PASS
443	442	15	24	19.3	35968	PASS

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100618\
Data File : BN002956.D
Acq On : 05 Oct 2018 07:59
Operator : MJ/SJ
Sample : DFTPP31
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP31

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091418MA.M
Title : SVOA CALIBRATION
Last Update : Mon Oct 08 14:32:35 2018



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2455

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.9	75474	PASS
68	69	0.00	2	1.8	1462	PASS
69	198	0.00	100	37.2	80458	PASS
70	69	0.00	2	0.6	445	PASS
127	198	10	80	45.2	97586	PASS
197	198	0.00	2	0.5	1000	PASS
198	198	100	100	100.0	216128	PASS
199	198	5	9	6.9	15009	PASS
275	198	10	60	26.9	58146	PASS
365	198	1	100	3.1	6687	PASS
441	198	0.01	100	13.1	28271	PASS
442	198	50	100	86.4	186712	PASS
443	442	15	24	19.3	35968	PASS

m/z	Abundance
38.00	641.0
39.00	4625.0
40.00	558.0
41.00	358.0
44.00	1852.0
49.00	697.0
50.00	14914.0
51.00	62608.0
52.00	3886.0
55.00	406.0
56.00	2112.0
56.90	4864.0
61.00	1083.0
62.00	1094.0
63.00	2757.0
63.90	446.0
65.00	1501.0
68.00	1357.0
69.00	66232.0
70.00	335.0
73.00	751.0
74.00	6580.0
75.00	10426.0
76.10	3298.0
77.00	73880.0
78.00	5161.0
79.00	4984.0
80.00	3558.0
81.00	5208.0
81.90	1310.0
83.00	1324.0
84.90	735.0
85.90	1319.0
87.00	591.0
90.90	1232.0
91.90	1201.0
93.00	8107.0
93.90	502.0
95.90	465.0
98.00	5818.0
98.90	4650.0
99.90	549.0
101.00	2829.0
102.90	1230.0
104.00	1653.0
105.00	1628.0
106.00	622.0
107.00	20496.0
108.00	3562.0
109.00	550.0
110.00	39344.0
110.90	6220.0
112.00	922.0
116.00	1049.0
117.00	13858.0
117.90	1289.0
122.00	1269.0
123.00	2286.0
123.90	1027.0
124.90	766.0
126.10	398.0
127.00	76376.0
128.00	6061.0
129.00	28192.0
130.00	2208.0
131.00	618.0
133.90	884.0
135.00	2419.0
135.90	1041.0
137.00	1304.0
139.90	434.0
141.00	3441.0
142.00	1237.0
142.90	869.0
146.10	505.0
147.00	1998.0
148.00	3982.0
149.00	855.0
151.30	586.0
152.90	1095.0
154.00	826.0
154.90	2250.0
156.00	3478.0
156.90	559.0
157.80	864.0
159.10	592.0
159.90	1061.0
161.00	1834.0
161.90	589.0
165.00	1165.0
166.00	1056.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

167.00	6253.0
167.90	3705.0
168.90	756.0
171.10	358.0
171.90	586.0
172.90	809.0
174.00	1392.0
174.90	2919.0
175.80	764.0
176.20	660.0
176.90	1314.0
177.90	311.0
178.90	5323.0
179.90	3691.0
181.00	1725.0
183.80	539.0
185.00	2542.0
186.00	20456.0
187.00	5416.0
188.00	852.0
189.00	1151.0
191.00	723.0
192.00	1793.0
193.00	2185.0
194.00	583.0
195.90	4730.0
197.90	154496.0
198.90	10284.0
200.00	751.0
201.40	774.0
202.90	1016.0
204.00	4960.0
205.00	8582.0
206.00	38680.0
207.00	5892.0
208.00	1368.0
209.00	502.0
210.20	741.0
211.10	1302.0
211.70	423.0
215.00	551.0
216.00	891.0
216.90	9911.0
217.90	1336.0
221.00	6243.0
223.00	1862.0
224.00	19616.0
225.00	5249.0
225.90	800.0
227.00	6882.0
228.00	1222.0
229.00	1738.0
230.90	865.0
233.80	362.0
234.80	514.0
236.10	359.0
237.00	845.0
238.90	471.0
240.90	551.0
241.90	1164.0
243.00	1041.0
244.00	16190.0
245.00	2423.0
246.00	2571.0
246.90	594.0
249.00	647.0
254.10	680.0
255.00	80584.0
256.00	13062.0
257.00	982.0
258.00	3817.0
258.90	932.0
264.90	1789.0
265.90	322.0
272.90	2372.0
274.00	6531.0
275.00	37424.0
276.00	5445.0
276.90	2802.0
277.90	538.0
281.10	336.0
284.90	538.0
292.90	650.0
296.00	9044.0
296.90	1360.0
303.10	912.0
313.90	310.0
315.00	967.0
316.00	644.0
323.00	3605.0
324.10	588.0
326.90	627.0
328.00	351.0
332.90	310.0
334.00	2180.0

Instrument :

BNA_N

ClientSampleId :

DFTPP31

335.00	440.0
341.00	321.0
345.90	552.0
351.90	691.0
353.00	591.0
354.00	818.0
355.20	374.0
364.90	3676.0
366.00	642.0
370.80	306.0
372.00	1740.0
373.00	424.0
382.90	427.0
390.10	356.0
401.90	634.0
403.00	815.0
403.90	398.0
420.90	715.0
422.00	777.0
423.00	5431.0
424.00	1671.0
441.00	14543.0
442.00	97160.0
443.00	19208.0
444.00	1884.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

m/z	Abundance
38.00	1067.0
39.10	5832.0
39.90	829.0
41.00	405.0
44.00	1698.0
49.00	421.0
50.00	21760.0
51.00	90376.0
52.00	4518.0
54.90	582.0
56.00	2864.0
57.00	6728.0
61.00	1105.0
61.90	1073.0
63.00	3900.0
64.00	581.0
65.00	1861.0
68.10	1450.0
69.00	97424.0
70.00	694.0
73.00	1187.0
74.00	9217.0
75.00	14315.0
76.00	5025.0
77.00	106032.0
78.00	7613.0
79.00	6930.0
80.00	5605.0
81.00	7317.0
82.00	2058.0
82.90	1830.0
84.90	1468.0
86.00	1882.0
87.10	904.0
88.00	516.0
91.00	1476.0
91.90	1937.0
93.00	11290.0
93.90	984.0
94.90	335.0
96.00	458.0
98.00	8362.0
99.00	7029.0
100.00	517.0
101.00	4304.0
101.90	340.0
103.10	1549.0
103.90	2740.0
105.00	2427.0
106.00	712.0
107.00	30744.0
108.00	5575.0
109.00	1198.0
110.00	61432.0
111.00	8714.0
111.90	1122.0
115.90	1997.0
117.00	21144.0
118.00	1948.0
119.80	443.0
121.90	2031.0
122.90	3087.0
123.90	1310.0
124.90	1567.0
127.00	116064.0
128.00	8904.0
129.00	43848.0
130.00	3323.0
131.00	699.0
132.00	303.0
133.10	421.0
134.00	1321.0
134.90	3662.0
135.90	1294.0
137.00	1844.0
137.90	370.0
139.90	540.0
140.90	5474.0
141.90	1943.0
142.90	1288.0
144.90	573.0
146.00	825.0
146.90	2966.0
148.00	5729.0
148.90	1380.0
150.00	506.0
151.00	769.0
151.70	571.0
152.90	1596.0
154.00	1298.0
155.00	3416.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

156.00	4580.0
157.10	818.0
158.00	1357.0
159.00	540.0
159.90	1888.0
161.00	2497.0
161.90	847.0
163.80	336.0
164.90	2235.0
166.00	1864.0
167.00	10686.0
167.90	5410.0
168.90	814.0
169.90	319.0
170.80	542.0
172.00	845.0
173.00	1396.0
174.00	2112.0
175.00	4425.0
175.90	1162.0
176.90	2395.0
178.00	879.0
178.90	8201.0
180.00	6212.0
181.00	2704.0
181.90	474.0
182.90	380.0
184.00	660.0
185.00	4392.0
186.00	30936.0
187.00	9649.0
188.00	1134.0
188.90	2358.0
189.90	358.0
191.00	892.0
192.00	2867.0
193.00	3231.0
194.00	681.0
195.10	507.0
196.00	7023.0
196.70	1613.0
197.90	253312.0
198.90	17944.0
199.90	1421.0
201.40	1165.0
203.00	1690.0
204.00	8621.0
205.00	13391.0
206.00	63856.0
207.00	9618.0
208.00	2214.0
208.90	779.0
210.90	2721.0
215.00	583.0
216.00	1208.0
216.90	16776.0
217.90	1690.0
221.00	11807.0
221.80	2259.0
222.90	3767.0
224.00	34760.0
225.00	8751.0
226.00	1342.0
227.00	13801.0
228.00	2111.0
229.00	2713.0
229.90	439.0
230.90	1418.0
234.00	980.0
235.00	1002.0
235.80	605.0
236.90	1214.0
238.90	631.0
240.00	373.0
240.90	835.0
242.00	1628.0
243.00	1739.0
244.00	27664.0
245.00	3531.0
246.00	5354.0
247.00	899.0
248.90	1105.0
251.00	419.0
252.90	581.0
253.90	907.0
255.00	143616.0
256.00	20920.0
257.00	1747.0
258.00	7882.0
259.00	1234.0
264.90	3186.0
265.90	419.0
270.90	502.0
271.90	327.0
273.00	4604.0

Instrument :

BNA_N

ClientSampleId :

DFTPP31

274.00	11548.0
275.00	68936.0
276.00	8848.0
277.00	4897.0
277.90	878.0
281.90	353.0
283.00	529.0
283.90	459.0
285.00	1025.0
289.00	352.0
292.90	1185.0
295.00	438.0
296.00	15868.0
296.90	2206.0
300.90	372.0
303.00	2231.0
303.90	571.0
309.90	340.0
314.00	1036.0
314.90	1988.0
316.00	1322.0
320.90	401.0
322.00	342.0
323.00	5748.0
323.90	1109.0
326.80	1160.0
327.90	473.0
332.10	362.0
332.80	759.0
334.00	3279.0
335.00	1206.0
340.90	980.0
346.00	1336.0
352.00	1612.0
353.00	1434.0
354.00	1825.0
355.10	396.0
365.00	7883.0
365.90	1084.0
371.10	460.0
372.00	3170.0
372.90	773.0
383.00	781.0
384.00	469.0
390.00	519.0
391.10	446.0
401.90	1558.0
402.90	1997.0
404.10	797.0
421.00	1759.0
422.00	1475.0
423.00	11060.0
424.00	2913.0
424.90	454.0
441.00	32120.0
442.00	211456.0
443.00	41736.0
444.00	3632.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

m/z	Abundance
37.00	331.0
38.00	759.0
39.00	4904.0
39.90	678.0
41.00	318.0
44.00	1879.0
48.90	469.0
50.00	18744.0
51.00	73440.0
52.00	3667.0
55.00	382.0
56.00	2225.0
57.00	5026.0
60.90	971.0
62.00	878.0
63.00	3143.0
64.10	499.0
65.00	1830.0
68.10	1579.0
69.00	77720.0
70.10	307.0
73.10	962.0
74.00	7779.0
75.00	12193.0
76.00	3745.0
77.00	87512.0
78.00	6104.0
79.00	6038.0
79.90	4109.0
81.00	6526.0
81.90	1592.0
83.00	1537.0
85.00	1295.0
86.00	1807.0
86.90	651.0
87.90	395.0
90.90	1397.0
91.90	1248.0
93.00	9677.0
94.00	600.0
95.90	567.0
98.00	7073.0
99.00	5447.0
99.90	404.0
101.00	3432.0
102.90	1338.0
104.00	2169.0
104.90	1956.0
106.00	949.0
107.00	26896.0
108.00	4582.0
109.10	1013.0
110.00	52696.0
111.00	7873.0
112.00	991.0
113.00	326.0
116.00	1453.0
117.00	20400.0
118.00	1347.0
120.00	375.0
121.00	303.0
122.00	1963.0
123.00	3200.0
123.90	1184.0
125.00	1090.0
127.00	100320.0
128.00	7236.0
129.00	37160.0
130.00	3287.0
130.90	495.0
134.00	951.0
135.00	2992.0
136.00	1158.0
136.90	1388.0
137.80	429.0
139.90	600.0
141.00	4598.0
142.00	1517.0
142.90	1006.0
145.10	306.0
146.00	891.0
146.90	2675.0
148.00	5415.0
148.90	1171.0
151.20	848.0
152.90	1350.0
154.00	1355.0
155.00	2994.0
156.00	3954.0
157.00	1059.0
158.00	1113.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

158.80	701.0
160.00	1428.0
160.90	2400.0
161.90	563.0
164.00	303.0
165.00	1720.0
165.90	1545.0
167.00	9925.0
168.00	4707.0
168.90	935.0
171.00	578.0
172.00	740.0
172.90	1334.0
174.00	2236.0
175.00	4126.0
176.00	1543.0
177.00	2179.0
177.90	843.0
178.90	7768.0
180.00	5733.0
180.90	2532.0
181.90	416.0
182.90	316.0
184.00	662.0
184.90	2972.0
186.00	29432.0
187.00	7619.0
188.00	862.0
188.90	1962.0
190.10	398.0
191.00	1082.0
192.00	2720.0
192.90	3031.0
193.90	716.0
195.00	353.0
196.00	6828.0
196.70	1389.0
197.90	240576.0
198.90	16800.0
200.00	1234.0
201.50	1300.0
202.90	1473.0
204.00	7947.0
205.00	12745.0
206.00	60160.0
207.00	8676.0
207.90	2034.0
208.90	595.0
210.20	1210.0
210.80	2093.0
211.80	469.0
214.90	662.0
216.00	1436.0
216.90	14556.0
218.00	1786.0
221.00	10845.0
223.00	3542.0
224.00	33336.0
225.00	7755.0
226.00	908.0
227.00	13209.0
227.90	2028.0
229.00	2788.0
229.80	319.0
231.00	1197.0
233.90	937.0
235.00	920.0
235.90	578.0
237.00	1029.0
239.00	447.0
239.90	443.0
240.90	589.0
242.00	1950.0
243.10	1687.0
244.00	27512.0
245.00	3254.0
246.00	4813.0
247.00	663.0
248.90	937.0
253.00	637.0
255.00	140928.0
256.00	19248.0
257.00	1629.0
258.00	7595.0
259.00	898.0
265.00	3105.0
265.90	341.0
271.00	328.0
273.00	4754.0
274.00	11720.0
275.00	68080.0
276.00	8709.0
277.00	4702.0
277.90	929.0
282.90	657.0

Instrument :

BNA_N

ClientSampleId :

DFTPP31

283.90	523.0
284.90	988.0
292.90	1151.0
295.10	410.0
295.90	17664.0
297.00	2543.0
303.00	2258.0
303.90	605.0
309.90	316.0
313.90	754.0
314.90	2238.0
315.90	1030.0
321.00	584.0
321.90	302.0
323.00	6535.0
324.00	1005.0
326.90	864.0
327.90	557.0
331.90	683.0
332.90	542.0
333.90	4024.0
335.00	921.0
341.00	612.0
346.00	1163.0
352.00	1844.0
353.00	1190.0
354.00	2051.0
355.00	522.0
364.90	8502.0
366.00	1391.0
370.90	629.0
372.00	3634.0
373.00	976.0
382.90	1216.0
390.00	454.0
391.00	343.0
392.00	387.0
402.00	1373.0
403.00	1985.0
404.00	996.0
421.00	1882.0
422.00	1968.0
423.00	14193.0
423.90	3135.0
424.90	393.0
441.00	38152.0
442.00	251520.0
443.00	46960.0
444.00	4491.0
444.90	349.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

Scan 2455 (17.428 min): BN002956.D
DFTPP31

m/z	Abundance
40.00	360.0
44.00	1822.0
73.00	359.0
207.10	525.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31