

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN100719\
 Data File : BN008050.D
 Acq On : 07 Oct 2019 15:11
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 23:45:12 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 03 18:39:49 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	36#	-0.03
2	1,4-Dioxane	0.446	0.492	-10.3	47#	-0.04
3	n-Nitrosodimethylamine	0.204	0.000	100.0#	0#	-3.01#
4 S	2-Fluorophenol	1.377	1.248	9.4	38#	-0.02
5 S	Phenol-d6	1.962	1.587	19.1	37#	0.00
6	bis(2-Chloroethyl)ether	1.133	1.329	-17.3	48#	-0.04
7 I	Naphthalene-d8	1.000	1.000	0.0	37#	-0.03
8 S	Nitrobenzene-d5	0.350	0.347	0.9	38#	-0.03
9	Naphthalene	1.123	1.148	-2.2	38#	-0.04
10	Hexachlorobutadiene	0.236	0.265	-12.3	42#	-0.04
11 SURR	2-Methylnaphthalene-d10	0.674	0.713	-5.8	41#	-0.03
12	2-Methylnaphthalene	0.759	0.802	-5.7	41#	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	39#	-0.03
14 S	2,4,6-Tribromophenol	0.218	0.217	0.5	46#	-0.02
15 S	2-Fluorobiphenyl	1.647	1.500	8.9	38#	-0.02
16	Acenaphthylene	2.121	1.987	6.3	40#	-0.03
17	Acenaphthene	1.368	1.276	6.7	40#	-0.03
18	Fluorene	1.753	1.635	6.7	39#	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	38#	-0.02
20	4-Bromophenyl-phenylether	0.243	0.234	3.7	37#	-0.03
21	Hexachlorobenzene	0.265	0.278	-4.9	40#	-0.03
22	Pentachlorophenol	0.098	0.113	-15.3	50	-0.02
23	Phenanthrene	1.246	1.283	-3.0	40#	-0.03
24	Anthracene	1.125	1.145	-1.8	40#	-0.02
25 SURR	Fluoranthene-d10	1.348	1.428	-5.9	41#	-0.03
26	Fluoranthene	1.549	1.719	-11.0	43#	-0.03
27 I	Chrysene-d12	1.000	1.000	0.0	53	-0.02
28	Pvrene	1.364	1.094	19.8	44#	-0.03
29 S	Terphenyl-d14	0.983	0.885	10.0	50#	-0.03
30	Benzo(a)anthracene	1.473	1.178	20.0	44#	-0.02
31	Chrysene	1.436	1.278	11.0	49#	-0.02
32	Bis(2-ethylhexyl)phthalate	0.758	1.355	-78.8#	106	-0.04
33	Indeno(1,2,3-cd)pyrene	1.797	1.640	8.7	50#	-0.05
34 I	Perylene-d12	1.000	1.000	0.0	47#	-0.04
35	Benzo(b)fluoranthene	1.389	1.399	-0.7	51	-0.03
36	Benzo(k)fluoranthene	1.373	1.342	2.3	48#	-0.03
37 C	Benzo(a)pyrene	1.305	1.256	3.8	48#	-0.03
38	Dibenzo(a,h)anthracene	1.335	1.347	-0.9	50#	-0.05
39	Benzo(a,h,i)perylene	1.321	1.360	-3.0	51	-0.05

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN100719\
Data File : BN008050.D
Acq On : 07 Oct 2019 15:11
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 07 23:45:12 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 03 18:39:49 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			